

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020720\
 Data File : VX014873.D
 Acq On : 07 Feb 2020 16:00
 Operator : JC/SP
 Sample : L1434-01 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EFF-WW

Quant Time: Feb 07 22:53:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jan 15 12:41:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	65966	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	382103	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	349337	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.06	65	149078	32.15	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	107.17%
60) 4-Bromofluorobenzene	11.13	95	167217	29.91	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	99.70%
63) Toluene-d8	8.71	98	454268	28.79	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	95.97%

Target Compounds

					Ovalue	
15) Acetone	2.43	58	1520993	2009.963	ug/l	87
20) Diisopropyl ether	3.84	45	81137	6.068	ug/l #	84
58) 4-Methyl-2-Pentanone	8.64	43	33757	5.131	ug/l	97
59) 2-Hexanone	9.49	43	7157	1.407	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020720\
Data File : VX014873.D
Acq On : 07 Feb 2020 16:00
Operator : JC/SP
Sample : L1434-01 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EFF-WW

Quant Time: Feb 07 22:53:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jan 15 12:41:12 2020
Response via : Initial Calibration

