

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020722\
 Data File : VX026786.D
 Acq On : 07 Feb 2022 15:52
 Operator : JC/MD
 Sample : N1359-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11R

Quant Time: Feb 07 23:30:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126769	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	210595	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190943	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82121	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80262	45.324	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.640%
35) Dibromofluoromethane	5.391	113	66056	48.572	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.140%
50) Toluene-d8	8.653	98	246509	50.301	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.600%
62) 4-Bromofluorobenzene	11.079	95	89958	51.673	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.340%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2107	2.333	ug/l	93
19) Methyl tert-butyl Ether	3.117	73	2197	0.457	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020722\
Data File : VX026786.D
Acq On : 07 Feb 2022 15:52
Operator : JC/MD
Sample : N1359-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11R

Quant Time: Feb 07 23:30:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

