

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020722\
 Data File : VX026788.D
 Acq On : 07 Feb 2022 16:38
 Operator : JC/MD
 Sample : N1359-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-19R

Quant Time: Feb 07 23:31:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	215162	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76902	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81368	45.114	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.220%
35) Dibromofluoromethane	5.391	113	66998	48.219	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.440%
50) Toluene-d8	8.653	98	244335	48.799	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.600%
62) 4-Bromofluorobenzene	11.079	95	85317	47.967	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.940%
Target Compounds						
16) Acetone	2.386	43	2254	2.451	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020722\
Data File : VX026788.D
Acq On : 07 Feb 2022 16:38
Operator : JC/MD
Sample : N1359-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-19R

Quant Time: Feb 07 23:31:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

