

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020722\
 Data File : VX026792.D
 Acq On : 07 Feb 2022 18:11
 Operator : JC/MD
 Sample : PB142502TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142502TB

Quant Time: Feb 07 23:32:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	129047	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	214844	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186266	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	79169	43.917	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.840%
35) Dibromofluoromethane	5.397	113	66129	47.664	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.320%
50) Toluene-d8	8.652	98	245724	49.149	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.079	95	88233	49.680	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.360%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	4189	4.557	ug/l	96
43) Isopropyl Acetate	6.348	43	11732	3.278	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	7142	1.287	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020722\
Data File : VX026792.D
Acq On : 07 Feb 2022 18:11
Operator : JC/MD
Sample : PB142502TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142502TB

Quant Time: Feb 07 23:32:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

