

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020722\
 Data File : VX026794.D
 Acq On : 07 Feb 2022 18:57
 Operator : JC/MD
 Sample : N1400-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KTS1-WC-S-MH-01-02-0-15

Quant Time: Feb 07 23:33:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	133719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	221144	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85020	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	82119	43.962	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.920%
35) Dibromofluoromethane	5.391	113	68659	48.078	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.160%
50) Toluene-d8	8.653	98	253588	49.277	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.560%
62) 4-Bromofluorobenzene	11.079	95	91324	49.955	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.920%
Target Compounds						
16) Acetone	2.380	43	13427	14.095	ug/l	91
43) Isopropyl Acetate	6.342	43	11298	3.067	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020722\
Data File : VX026794.D
Acq On : 07 Feb 2022 18:57
Operator : JC/MD
Sample : N1400-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KTS1-WC-S-MH-01-02-0-15

Quant Time: Feb 07 23:33:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

