

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020722\
 Data File : VX026795.D
 Acq On : 07 Feb 2022 19:20
 Operator : JC/MD
 Sample : N1419-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GRIT-STOCKPILE

Quant Time: Feb 07 23:33:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132323	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218576	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81395	44.034	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.060%
35) Dibromofluoromethane	5.391	113	68163	48.291	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	8.653	98	253690	49.876	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.760%
62) 4-Bromofluorobenzene	11.079	95	91106	50.421	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.840%
Target Compounds						
					Qvalue	
13) Acrolein	2.239	56	3167	11.591	ug/l	99
16) Acetone	2.386	43	58152	61.691	ug/l	96
18) Methyl Acetate	2.709	43	2290	1.222	ug/l	94
20) Methylene Chloride	2.800	84	1588	1.020	ug/l	96
25) 2-Butanone	4.562	43	16615	13.554	ug/l	97
37) Ethyl Acetate	4.721	43	11252	4.992	ug/l	99
43) Isopropyl Acetate	6.348	43	8791	2.414	ug/l	97
52) Toluene	8.720	92	26703	7.070	ug/l	100
95) Naphthalene	13.780	128	8506	1.416	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020722\
Data File : VX026795.D
Acq On : 07 Feb 2022 19:20
Operator : JC/MD
Sample : N1419-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GRIT-STOCKPILE

Quant Time: Feb 07 23:33:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

