

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
 Data File : VX034026.D
 Acq On : 06 Feb 2023 14:44
 Operator : JC/MD
 Sample : VX0207MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0207MBL01

Quant Time: Feb 07 00:50:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	147489	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	205130	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182260	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86863	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67547	47.506	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.020%
35) Dibromofluoromethane	5.385	113	63761	48.840	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.680%
50) Toluene-d8	8.647	98	231952	49.313	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.620%
62) 4-Bromofluorobenzene	11.079	95	79628	47.399	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.800%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
Data File : VX034026.D
Acq On : 06 Feb 2023 14:44
Operator : JC/MD
Sample : VX0207MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0207MBL01

Quant Time: Feb 07 00:50:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

