

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX020723\  
 Data File : VX034034.D  
 Acq On : 06 Feb 2023 17:52  
 Operator : JC/MD  
 Sample : 01421-07  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 BP-VPB-RW10-DUP-20230201

Quant Time: Feb 07 00:54:59 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X020423W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 03 23:25:45 2023  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|-----------------------------|----------------|------|------------|---------|--------|----------|
| -----                       |                |      |            |         |        |          |
| Internal Standards          |                |      |            |         |        |          |
| 1) Pentafluorobenzene       | 5.550          | 168  | 145446     | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763          | 114  | 202762     | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055         | 117  | 181839     | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024         | 152  | 89138      | 50.000  | ug/l   | 0.00     |
|                             |                |      |            |         |        |          |
| System Monitoring Compounds |                |      |            |         |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958          | 65   | 67948      | 48.459  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 78 - 117 |      | Recovery = | 96.920% |        |          |
| 35) Dibromofluoromethane    | 5.385          | 113  | 64039      | 49.626  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 75 - 124 |      | Recovery = | 99.260% |        |          |
| 50) Toluene-d8              | 8.647          | 98   | 230719     | 49.624  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 92 - 112 |      | Recovery = | 99.240% |        |          |
| 62) 4-Bromofluorobenzene    | 11.079         | 95   | 81367      | 49.000  | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 83 - 123 |      | Recovery = | 98.000% |        |          |
|                             |                |      |            |         |        |          |
| Target Compounds            |                |      |            |         | Qvalue |          |
| 16) Acetone                 | 2.380          | 43   | 1192       | 2.013   | ug/l   | 98       |
| -----                       |                |      |            |         |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX020723\  
Data File : VX034034.D  
Acq On : 06 Feb 2023 17:52  
Operator : JC/MD  
Sample : 01421-07  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BP-VPB-RW10-DUP-20230201

Quant Time: Feb 07 00:54:59 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X020423W.M  
Quant Title : SW846 8260  
QLast Update : Fri Feb 03 23:25:45 2023  
Response via : Initial Calibration

