

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
 Data File : VX034041.D
 Acq On : 06 Feb 2023 20:34
 Operator : JC/MD
 Sample : 01416-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2023-2-2

Quant Time: Feb 07 00:58:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	137880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	192331	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65609	49.359	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.720%
35) Dibromofluoromethane	5.385	113	60661	49.557	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.120%
50) Toluene-d8	8.647	98	220039	49.893	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.780%
62) 4-Bromofluorobenzene	11.079	95	77604	49.269	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.540%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	15003	26.724	ug/l	90
20) Methylene Chloride	2.788	84	1588	1.236	ug/l #	94
25) 2-Butanone	4.562	43	2697	3.299	ug/l #	87
51) 4-Methyl-2-Pentanone	8.574	43	1191	0.779	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
Data File : VX034041.D
Acq On : 06 Feb 2023 20:34
Operator : JC/MD
Sample : 01416-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2023-2-2

Quant Time: Feb 07 00:58:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

