

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
 Data File : VX034043.D
 Acq On : 06 Feb 2023 21:21
 Operator : JC/MD
 Sample : 01442-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2023-2-3

Quant Time: Feb 07 00:58:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136068	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189567	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82833	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64102	48.867	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.740%
35) Dibromofluoromethane	5.385	113	59805	49.570	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.140%
50) Toluene-d8	8.647	98	214219	49.282	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.560%
62) 4-Bromofluorobenzene	11.079	95	75218	48.450	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.900%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1318	2.379	ug/l	92
20) Methylene Chloride	2.788	84	409	0.323	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
Data File : VX034043.D
Acq On : 06 Feb 2023 21:21
Operator : JC/MD
Sample : 01442-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2023-2-3

Quant Time: Feb 07 00:58:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

