

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
 Data File : VX034047.D
 Acq On : 06 Feb 2023 22:53
 Operator : JC/MD
 Sample : 01455-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 07 01:00:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	136891	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	192373	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84737	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65451	49.596	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.200%
35) Dibromofluoromethane	5.385	113	60573	49.475	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.940%
50) Toluene-d8	8.647	98	216258	49.025	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.060%
62) 4-Bromofluorobenzene	11.079	95	75894	48.173	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
Data File : VX034047.D
Acq On : 06 Feb 2023 22:53
Operator : JC/MD
Sample : 01455-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 07 01:00:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

