

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
 Data File : VX034048.D
 Acq On : 06 Feb 2023 23:16
 Operator : JC/MD
 Sample : 01455-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Feb 07 01:00:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	134779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	190384	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85066	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64324	49.505	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.020%
35) Dibromofluoromethane	5.385	113	59484	49.093	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.180%
50) Toluene-d8	8.647	98	215842	49.442	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.880%
62) 4-Bromofluorobenzene	11.079	95	77369	49.622	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
Data File : VX034048.D
Acq On : 06 Feb 2023 23:16
Operator : JC/MD
Sample : 01455-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Feb 07 01:00:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

