

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
 Data File : VX034051.D
 Acq On : 07 Feb 2023 00:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 07 02:53:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.406	0.397	2.2	84	0.00
3 P	Chloromethane	0.384	0.374	2.6	86	0.00
4 C	Vinyl Chloride	0.402	0.363	9.7#	80	0.00
5 T	Bromomethane	0.207	0.210	-1.4	92	-0.01
6 T	Chloroethane	0.222	0.214	3.6	84	0.00
7 T	Trichlorofluoromethane	0.639	0.618	3.3	85	0.00
8 T	Diethyl Ether	0.220	0.203	7.7	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.427	0.386	9.6	80	0.00
10 T	Methyl Iodide	0.605	0.582	3.8	83	0.00
11 T	Tert butyl alcohol	0.068	0.069	-1.5	95	-0.01
12 CM	1,1-Dichloroethene	0.409	0.379	7.3#	82	0.00
13 T	Acrolein	0.091	0.085	6.6	84	0.00
14 T	Allyl chloride	0.667	0.605	9.3	79	0.00
15 T	Acrylonitrile	0.217	0.202	6.9	83	0.00
16 T	Acetone	0.204	0.166	18.6	78	0.00
17 T	Carbon Disulfide	1.039	0.931	10.4	79	0.00
18 T	Methyl Acetate	0.633	0.598	5.5	82	0.00
19 T	Methyl tert-butyl Ether	1.262	1.195	5.3	83	0.00
20 T	Methylene Chloride	0.466	0.422	9.4	84	0.00
21 T	trans-1,2-Dichloroethene	0.446	0.411	7.8	83	0.00
22 T	Diisopropyl ether	1.317	1.212	8.0	82	0.00
23 T	Vinyl Acetate	1.003	0.936	6.7	80	0.00
24 P	1,1-Dichloroethane	0.759	0.706	7.0	83	0.00
25 T	2-Butanone	0.296	0.266	10.1	83	0.00
26 T	2,2-Dichloropropane	0.554	0.472	14.8	75	0.00
27 T	cis-1,2-Dichloroethene	0.513	0.476	7.2	83	0.00
28 T	Bromochloromethane	0.334	0.312	6.6	86	0.00
29 T	Tetrahydrofuran	0.189	0.174	7.9	83	0.00
30 C	Chloroform	0.791	0.753	4.8#	85	0.00
31 T	Cyclohexane	0.702	0.638	9.1	80	0.00
32 T	1,1,1-Trichloroethane	0.681	0.663	2.6	85	0.00
33 S	1,2-Dichloroethane-d4	0.482	0.442	8.3	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.318	0.293	7.9	87	0.00
36 T	1,1-Dichloropropene	0.436	0.397	8.9	82	0.00
37 T	Ethyl Acetate	0.406	0.376	7.4	81	0.00
38 T	Carbon Tetrachloride	0.465	0.441	5.2	85	0.00
39 T	Methylcyclohexane	0.549	0.498	9.3	81	0.00
40 TM	Benzene	1.262	1.176	6.8	83	0.00
41 T	Methacrylonitrile	0.229	0.213	7.0	84	0.00
42 TM	1,2-Dichloroethane	0.429	0.418	2.6	87	0.00
43 T	Isopropyl Acetate	0.634	0.601	5.2	82	0.00
44 TM	Trichloroethene	0.388	0.366	5.7	85	0.00
45 C	1,2-Dichloropropane	0.312	0.292	6.4#	83	0.00
46 T	Dibromomethane	0.223	0.217	2.7	87	0.00
47 T	Bromodichloromethane	0.423	0.414	2.1	85	0.00
48 T	Methyl methacrylate	0.327	0.315	3.7	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
 Data File : VX034051.D
 Acq On : 07 Feb 2023 00:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 07 02:53:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 03 23:25:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	98	0.00
50 S	Toluene-d8	1.147	1.049	8.5	87	0.00
51 T	4-Methyl-2-Pentanone	0.398	0.387	2.8	85	0.00
52 CM	Toluene	0.818	0.775	5.3#	85	0.00
53 T	t-1,3-Dichloropropene	0.425	0.418	1.6	83	0.00
54 T	cis-1,3-Dichloropropene	0.484	0.460	5.0	81	0.00
55 T	1,1,2-Trichloroethane	0.327	0.318	2.8	87	0.00
56 T	Ethyl methacrylate	0.456	0.459	-0.7	85	0.00
57 T	1,3-Dichloropropane	0.534	0.513	3.9	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.212	0.201	5.2	83	0.00
59 T	2-Hexanone	0.291	0.289	0.7	87	0.00
60 T	Dibromochloromethane	0.362	0.367	-1.4	88	0.00
61 T	1,2-Dibromoethane	0.351	0.340	3.1	86	0.00
62 S	4-Bromofluorobenzene	0.409	0.399	2.4	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.417	0.376	9.8	85	0.00
65 PM	Chlorobenzene	1.033	0.985	4.6	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.390	0.377	3.3	87	0.00
67 C	Ethyl Benzene	1.748	1.679	3.9#	88	0.00
68 T	m/p-Xylenes	0.699	0.670	4.1	88	0.00
69 T	o-Xylene	0.671	0.653	2.7	88	0.00
70 T	Styrene	1.105	1.097	0.7	89	0.00
71 P	Bromoform	0.318	0.329	-3.5	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.342	3.167	5.2	89	0.00
74 T	N-amyl acetate	1.101	1.087	1.3	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.005	0.938	6.7	91	0.00
76 T	1,2,3-Trichloropropane	0.859	0.856	0.3	102	0.00
77 T	Bromobenzene	0.927	0.884	4.6	91	0.00
78 T	n-propylbenzene	3.695	3.488	5.6	88	0.00
79 T	2-Chlorotoluene	2.237	2.135	4.6	91	0.00
80 T	1,3,5-Trimethylbenzene	2.781	2.666	4.1	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.279	0.254	9.0	84	0.00
82 T	4-Chlorotoluene	2.530	2.407	4.9	90	0.00
83 T	tert-Butylbenzene	2.874	2.743	4.6	91	0.00
84 T	1,2,4-Trimethylbenzene	2.779	2.698	2.9	92	0.00
85 T	sec-Butylbenzene	3.438	3.257	5.3	88	0.00
86 T	p-Isopropyltoluene	3.015	2.894	4.0	90	0.00
87 T	1,3-Dichlorobenzene	1.700	1.603	5.7	90	0.00
88 T	1,4-Dichlorobenzene	1.726	1.603	7.1	91	0.00
89 T	n-Butylbenzene	2.364	2.255	4.6	88	0.00
90 T	Hexachloroethane	0.475	0.466	1.9	91	0.00
91 T	1,2-Dichlorobenzene	1.658	1.568	5.4	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.198	0.194	2.0	90	0.00
93 T	1,2,4-Trichlorobenzene	1.172	1.128	3.8	90	0.00
94 T	Hexachlorobutadiene	0.580	0.521	10.2	90	0.00
95 T	Naphthalene	3.219	3.175	1.4	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020723\
Data File : VX034051.D
Acq On : 07 Feb 2023 00:26
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 07 02:53:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020423W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 03 23:25:45 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.177	1.099	6.6	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6