

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020724\
 Data File : VX040148.D
 Acq On : 07 Feb 2024 12:46
 Operator : JC/MD
 Sample : VX0207WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0207WBL01

Quant Time: Feb 07 15:22:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	106919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189692	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	181315	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90920	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75329	43.777	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.560%
35) Dibromofluoromethane	5.385	113	58953	45.654	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.300%
50) Toluene-d8	8.647	98	232540	48.154	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.300%
62) 4-Bromofluorobenzene	11.079	95	89886	48.340	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020724\
Data File : VX040148.D
Acq On : 07 Feb 2024 12:46
Operator : JC/MD
Sample : VX0207WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0207WBL01

Quant Time: Feb 07 15:22:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

