

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020724\
 Data File : VX040150.D
 Acq On : 07 Feb 2024 14:32
 Operator : JC/MD
 Sample : P1376-03RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-740-742RE

Quant Time: Feb 07 15:23:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	83960	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	139168	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	144601	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35955	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	44727	33.101	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	66.200%#
35) Dibromofluoromethane	5.385	113	34911	36.851	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	73.700%#
50) Toluene-d8	8.647	98	140827	39.749	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	79.500%#
62) 4-Bromofluorobenzene	11.079	95	38536	28.248	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	56.500%#
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	1732	2.736	ug/l	92
17) Carbon Disulfide	2.514	76	1882	0.682	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020724\
Data File : VX040150.D
Acq On : 07 Feb 2024 14:32
Operator : JC/MD
Sample : P1376-03RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-740-742RE

Quant Time: Feb 07 15:23:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

