

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020724\
 Data File : VX040154.D
 Acq On : 07 Feb 2024 16:05
 Operator : JC/MD
 Sample : P1337-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1

Quant Time: Feb 08 00:35:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John Carlone 02/08/2024
 Supervised By :Semsettin Yesilyurt 02/08/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	52350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	106280	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105357	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42435	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	47669	56.580	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 113.160%			
35) Dibromofluoromethane	5.385	113	33463	46.253	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 92.500%			
50) Toluene-d8	8.647	98	137344	50.762	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.520%			
62) 4-Bromofluorobenzene	11.079	95	51547	49.478	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.960%			
Target Compounds						
						Qvalue
16) Acetone	2.380	43	28685	72.665	ug/l	98
20) Methylene Chloride	2.788	84	3626	5.169	ug/l	97
25) 2-Butanone	4.586	43	3994m	7.278	ug/l	
43) Isopropyl Acetate	6.348	43	52239	28.989	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020724\
Data File : VX040154.D
Acq On : 07 Feb 2024 16:05
Operator : JC/MD
Sample : P1337-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1

Quant Time: Feb 08 00:35:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 02/08/2024
Supervised By : Semsettin Yesilyurt 02/08/2024

