

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020724\
 Data File : VX040158.D
 Acq On : 07 Feb 2024 17:37
 Operator : JC/MD
 Sample : P1337-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6

Quant Time: Feb 08 00:35:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	49906	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	95284	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49365	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	45450	56.588	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.180%	
35) Dibromofluoromethane	5.391	113	31397	48.405	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.820%	
50) Toluene-d8	8.647	98	126463	52.135	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.260%	
62) 4-Bromofluorobenzene	11.079	95	48678	52.117	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.240%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	29835	79.280	ug/l	97
20) Methylene Chloride	2.788	84	5159	7.714	ug/l	92
25) 2-Butanone	4.581	43	3204	6.125	ug/l	94
43) Isopropyl Acetate	6.342	43	43869	27.153	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020724\
Data File : VX040158.D
Acq On : 07 Feb 2024 17:37
Operator : JC/MD
Sample : P1337-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6

Quant Time: Feb 08 00:35:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

