

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020724\
 Data File : VX040159.D
 Acq On : 07 Feb 2024 18:00
 Operator : JC/MD
 Sample : P1141-12RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2-0-2DRE

Quant Time: Feb 08 00:35:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 01 05:44:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	45782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	91631	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	91783	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	29626	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	45313	61.499	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 123.000%#			
35) Dibromofluoromethane	5.391	113	30330	48.624	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.240%			
50) Toluene-d8	8.647	98	121414	52.049	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 104.100%			
62) 4-Bromofluorobenzene	11.079	95	45814	51.006	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.020%			
Target Compounds						
16) Acetone	2.380	43	7802	22.600	ug/l	98
20) Methylene Chloride	2.788	84	2684	4.375	ug/l	91
43) Isopropyl Acetate	6.342	43	55638	35.811	ug/l	96
52) Toluene	8.720	92	16134	9.852	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020724\
Data File : VX040159.D
Acq On : 07 Feb 2024 18:00
Operator : JC/MD
Sample : P1141-12RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2-0-2DRE

Quant Time: Feb 08 00:35:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X013024W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 01 05:44:39 2024
Response via : Initial Calibration

