

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020725\
 Data File : VX044861.D
 Acq On : 07 Feb 2025 12:16
 Operator : JC/MD
 Sample : Q1319-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 002-35TH-AVE(JAN)

Quant Time: Feb 07 12:41:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011625W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Jan 17 01:21:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	17858	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	100352	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.049	117	88457	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.946	65	49084	28.471	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	94.900%
60) 4-Bromofluorobenzene	11.079	95	48085	27.987	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.300%
63) Toluene-d8	8.647	98	146971	30.548	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.833%
Target Compounds						
					Qvalue	
15) Acetone	2.380	58	5137	27.282	ug/l	92
25) Chloroform	5.093	83	1751	0.688	ug/l #	79
62) Toluene	8.714	91	90313	15.505	ug/l	97
82) p-Isopropyltoluene	12.006	119	3864	0.799	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020725\
Data File : VX044861.D
Acq On : 07 Feb 2025 12:16
Operator : JC/MD
Sample : Q1319-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
002-35TH-AVE(JAN)

Quant Time: Feb 07 12:41:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011625W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Jan 17 01:21:41 2025
Response via : Initial Calibration

