

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020819\
 Data File : VX007442.D
 Acq On : 08 Feb 2019 15:16
 Operator : JC/SP
 Sample : K1441-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SEWER-OUTLET-02

Quant Time: Feb 08 23:44:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X012519W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 28 10:35:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	77865	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	417503	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	369266	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.07	65	157866	30.89	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.97%
60) 4-Bromofluorobenzene	11.13	95	161336	28.31	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	94.37%
63) Toluene-d8	8.71	98	486857	29.25	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	97.50%

Target Compounds						Ovalue
15) Acetone	2.45	58	2208018	3031.629	ug/l	91
25) Chloroform	5.21	83	40194	5.234	ug/l	96
41) Bromodichloromethane	7.90	83	13269	2.493	ug/l #	96
53) Dibromochloromethane	9.59	129	3629	0.789	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020819\
Data File : VX007442.D
Acq On : 08 Feb 2019 15:16
Operator : JC/SP
Sample : K1441-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SEWER-OUTLET-02

Quant Time: Feb 08 23:44:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X012519W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 28 10:35:22 2019
Response via : Initial Calibration

