

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020819\
 Data File : VX007447.D
 Acq On : 08 Feb 2019 17:29
 Operator : JC/SP
 Sample : K1445-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 11 09:39:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	508018	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	793570	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	741943	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	360080	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	277696	51.50	ug/l	0.00
Spiked Amount			Recovery	=	103.00%	
35) Dibromofluoromethane	5.51	113	252928	48.98	ug/l	0.00
Spiked Amount			Recovery	=	97.96%	
50) Toluene-d8	8.71	98	945473	49.35	ug/l	0.00
Spiked Amount			Recovery	=	98.70%	
62) 4-Bromofluorobenzene	11.13	95	335740	48.00	ug/l	0.00
Spiked Amount			Recovery	=	96.00%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020819\
Data File : VX007447.D
Acq On : 08 Feb 2019 17:29
Operator : JC/SP
Sample : K1445-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 11 09:39:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

