

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020819\
 Data File : VX007451.D
 Acq On : 08 Feb 2019 19:15
 Operator : JC/SP
 Sample : K1441-02DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SEWER-OUTLET-02DL

Quant Time: Feb 11 09:35:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X012519W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 28 10:35:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	76820	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	401579	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	354985	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	151918	30.13	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.43%
60) 4-Bromofluorobenzene	11.13	95	153950	28.10	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	93.67%
63) Toluene-d8	8.71	98	468586	29.29	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	97.63%

Target Compounds					Ovalue
15) Acetone	2.45	58	194497	270.679 ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020819\
Data File : VX007451.D
Acq On : 08 Feb 2019 19:15
Operator : JC/SP
Sample : K1441-02DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SEWER-OUTLET-02DL

Quant Time: Feb 11 09:35:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X012519W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 28 10:35:22 2019
Response via : Initial Calibration

