

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX020819\
 Data File : VX007433.D
 Acq On : 08 Feb 2019 10:39
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 08 23:15:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	114	0.00
2 T	Dichlorodifluoromethane	0.557	0.568	-2.0	103	0.00
3 P	Chloromethane	0.563	0.482	14.4	100	0.00
4 C	Vinyl Chloride	0.552	0.513	7.1#	103	0.00
5 T	Bromomethane	0.629	0.512	18.6	105	0.00
6 T	Chloroethane	0.357	0.322	9.8	108	0.00
7 T	Trichlorofluoromethane	0.864	0.806	6.7	109	0.00
8 T	Diethyl Ether	0.325	0.291	10.5	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.514	0.476	7.4	112	0.00
10 T	Methyl Iodide	0.756	0.691	8.6	102	0.00
11 T	Tert butyl alcohol	0.110	0.099	10.0	112	0.00
12 CM	1,1-Dichloroethene	0.477	0.422	11.5#	107	0.00
13 T	Acrolein	0.052	0.050	3.8	109	0.00
14 T	Allyl chloride	0.778	0.720	7.5	108	0.00
15 T	Acrylonitrile	0.272	0.248	8.8	114	0.00
16 T	Acetone	0.243	0.221	9.1	112	0.00
17 T	Carbon Disulfide	1.260	1.056	16.2	103	0.00
18 T	Methyl Acetate	0.688	0.641	6.8	120	0.00
19 T	Methyl tert-butyl Ether	1.547	1.406	9.1	111	0.00
20 T	Methylene Chloride	0.550	0.464	15.6	108	0.00
21 T	trans-1,2-Dichloroethene	0.509	0.450	11.6	108	0.00
22 T	Diisopropyl ether	1.431	1.377	3.8	115	0.00
23 T	Vinyl Acetate	1.187	1.200	-1.1	117	0.00
24 P	1,1-Dichloroethane	0.822	0.794	3.4	115	0.00
25 T	2-Butanone	0.346	0.318	8.1	113	0.00
26 T	2,2-Dichloropropane	0.626	0.598	4.5	110	0.00
27 T	cis-1,2-Dichloroethene	0.555	0.483	13.0	107	0.00
28 T	Bromochloromethane	0.378	0.396	-4.8	120	0.00
29 T	Tetrahydrofuran	0.218	0.208	4.6	115	0.00
30 C	Chloroform	0.857	0.785	8.4#	110	0.00
31 T	Cyclohexane	0.738	0.670	9.2	108	0.00
32 T	1,1,1-Trichloroethane	0.724	0.686	5.2	110	0.00
33 S	1,2-Dichloroethane-d4	0.531	0.571	-7.5	128	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.325	0.371	-14.2	128	0.00
36 T	1,1-Dichloropropene	0.449	0.411	8.5	107	0.00
37 T	Ethyl Acetate	0.453	0.425	6.2	112	0.00
38 T	Carbon Tetrachloride	0.435	0.435	0.0	109	0.00
39 T	Methylcyclohexane	0.543	0.538	0.9	112	0.00
40 TM	Benzene	1.321	1.246	5.7	109	0.00
41 T	Methacrylonitrile	0.243	0.239	1.6	112	0.00
42 TM	1,2-Dichloroethane	0.457	0.416	9.0	107	0.00
43 T	Isopropyl Acetate	0.700	0.685	2.1	111	0.00
44 TM	Trichloroethene	0.398	0.368	7.5	106	0.00
45 C	1,2-Dichloropropane	0.331	0.313	5.4#	109	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX020819\
 Data File : VX007433.D
 Acq On : 08 Feb 2019 10:39
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 08 23:15:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Feb 02 01:32:36 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.241	0.226	6.2	112	0.00
47 T	Bromodichloromethane	0.408	0.404	1.0	111	0.00
48 T	Methyl methacrylate	0.350	0.350	0.0	110	0.00
49 T	1,4-Dioxane	0.009	0.008	11.1	112	0.00
50 S	Toluene-d8	1.207	1.379	-14.3	126	0.00
51 T	4-Methyl-2-Pentanone	0.461	0.443	3.9	108	0.00
52 CM	Toluene	0.853	0.811	4.9#	108	0.00
53 T	t-1,3-Dichloropropene	0.438	0.455	-3.9	110	0.00
54 T	cis-1,3-Dichloropropene	0.482	0.497	-3.1	110	0.00
55 T	1,1,2-Trichloroethane	0.355	0.333	6.2	110	0.00
56 T	Ethyl methacrylate	0.492	0.474	3.7	106	0.00
57 T	1,3-Dichloropropane	0.565	0.533	5.7	109	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.238	8.1	100	0.00
59 T	2-Hexanone	0.359	0.340	5.3	109	0.00
60 T	Dibromochloromethane	0.336	0.355	-5.7	109	0.00
61 T	1,2-Dibromoethane	0.376	0.366	2.7	110	0.00
62 S	4-Bromofluorobenzene	0.441	0.504	-14.3	125	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.471	0.432	8.3	106	0.00
65 PM	Chlorobenzene	1.091	1.015	7.0	106	0.00
66 T	1,1,1,2-Tetrachloroethane	0.366	0.374	-2.2	111	0.00
67 C	Ethyl Benzene	1.757	1.716	2.3#	109	0.00
68 T	m/p-Xylenes	0.698	0.674	3.4	106	0.00
69 T	o-Xylene	0.672	0.651	3.1	106	0.00
70 T	Styrene	1.090	1.056	3.1	105	0.00
71 P	Bromoform	0.283	0.291	-2.8	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.398	3.289	3.2	108	0.00
74 T	N-amyl acetate	1.314	1.263	3.9	106	0.00
75 P	1,1,2,2-Tetrachloroethane	1.036	0.975	5.9	109	0.00
76 T	1,2,3-Trichloropropane	0.973	0.935	3.9	106	0.00
77 T	Bromobenzene	0.957	0.869	9.2	103	0.00
78 T	n-propylbenzene	3.726	3.736	-0.3	108	0.00
79 T	2-Chlorotoluene	2.261	2.130	5.8	105	0.00
80 T	1,3,5-Trimethylbenzene	2.780	2.686	3.4	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.280	0.297	-6.1	110	0.00
82 T	4-Chlorotoluene	2.616	2.497	4.5	104	0.00
83 T	tert-Butylbenzene	2.752	2.652	3.6	106	0.00
84 T	1,2,4-Trimethylbenzene	2.822	2.751	2.5	105	0.00
85 T	sec-Butylbenzene	3.331	3.342	-0.3	107	0.00
86 T	p-Isopropyltoluene	3.007	2.999	0.3	106	0.00
87 T	1,3-Dichlorobenzene	1.718	1.586	7.7	105	0.00
88 T	1,4-Dichlorobenzene	1.743	1.595	8.5	105	0.00
89 T	n-Butylbenzene	2.587	2.607	-0.8	107	0.00

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX020819\
Data File : VX007433.D
Acq On : 08 Feb 2019 10:39
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 08 23:15:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X020119W.M
Quant Title : SW846 8260
QLast Update : Sat Feb 02 01:32:36 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.450	0.481	-6.9	112	0.00
91 T	1,2-Dichlorobenzene	1.710	1.571	8.1	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.215	0.208	3.3	110	0.00
93 T	1,2,4-Trichlorobenzene	1.153	1.123	2.6	108	0.00
94 T	Hexachlorobutadiene	0.560	0.548	2.1	111	0.00
95 T	Naphthalene	3.372	3.338	1.0	109	0.00
96 T	1,2,3-Trichlorobenzene	1.161	1.107	4.7	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6