

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026815.D
 Acq On : 08 Feb 2022 12:31
 Operator : JC/MD
 Sample : VX0208WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 09 03:00:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	100530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	161727	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63097	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61656	51.092	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 102.180%			
35) Dibromofluoromethane	5.391	113	49666	51.546	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.100%			
50) Toluene-d8	8.653	98	180105	50.922	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.840%			
62) 4-Bromofluorobenzene	11.079	95	65424	52.540	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 105.080%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	21665	19.820	ug/l	97
3) Chloromethane	1.294	50	19006	18.690	ug/l	100
4) Vinyl Chloride	1.379	62	20180	18.957	ug/l	98
5) Bromomethane	1.617	94	9604	21.261	ug/l	95
6) Chloroethane	1.690	64	12109	18.967	ug/l	96
7) Trichlorofluoromethane	1.892	101	31327	19.998	ug/l	100
8) Diethyl Ether	2.135	74	11592	19.976	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.337	101	21077	20.526	ug/l	98
10) Methyl Iodide	2.459	142	22684	17.463	ug/l	97
11) Tert butyl alcohol	2.977	59	21829	100.130	ug/l	95
12) 1,1-Dichloroethene	2.324	96	19760	19.206	ug/l	94
13) Acrolein	2.239	56	21964	109.021	ug/l	99
14) Allyl chloride	2.672	41	36172	19.848	ug/l	97
15) Acrylonitrile	3.068	53	62496	100.038	ug/l	99
16) Acetone	2.385	43	59010	111.336	ug/l	95
17) Carbon Disulfide	2.520	76	55931	19.204	ug/l	98
18) Methyl Acetate	2.709	43	29068	19.061	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	72659	19.952	ug/l	100
20) Methylene Chloride	2.794	84	21972	19.132	ug/l	96
21) trans-1,2-Dichloroethene	3.099	96	21807	20.142	ug/l	96
22) Diisopropyl ether	3.763	45	74160	20.212	ug/l	95
23) Vinyl Acetate	3.727	43	316160	102.488	ug/l	98
24) 1,1-Dichloroethane	3.611	63	39411	19.850	ug/l	100
25) 2-Butanone	4.562	43	89231	104.492	ug/l	99
26) 2,2-Dichloropropane	4.483	77	29037	21.646	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	24944	20.067	ug/l	97
28) Bromochloromethane	4.903	49	16934	20.576	ug/l	93
29) Tetrahydrofuran	5.013	42	57733	99.447	ug/l	97
30) Chloroform	5.098	83	40540	19.653	ug/l	98
31) Cyclohexane	5.476	56	38850	20.261	ug/l	98
32) 1,1,1-Trichloroethane	5.391	97	35439	19.486	ug/l	99
36) 1,1-Dichloropropene	5.696	75	30558	20.103	ug/l	98
37) Ethyl Acetate	4.714	43	33736	19.529	ug/l	98
38) Carbon Tetrachloride	5.684	117	31763	19.927	ug/l	100
39) Methylcyclohexane	7.385	83	38504	20.655	ug/l	97
40) Benzene	6.043	78	88354	19.944	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026815.D
 Acq On : 08 Feb 2022 12:31
 Operator : JC/MD
 Sample : VX0208WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 09 03:00:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	18801	20.081	ug/l	97
42) 1,2-Dichloroethane	6.092	62	32549	20.850	ug/l	97
43) Isopropyl Acetate	6.348	43	54316	20.436	ug/l	99
44) Trichloroethene	7.128	130	26143	19.654	ug/l	95
45) 1,2-Dichloropropane	7.433	63	22444	20.024	ug/l	96
46) Dibromomethane	7.586	93	15667	20.200	ug/l	96
47) Bromodichloromethane	7.823	83	30688	20.662	ug/l	97
48) Methyl methacrylate	7.695	41	26703	20.276	ug/l	98
49) 1,4-Dioxane	7.677	88	10264	452.278	ug/l	94
51) 4-Methyl-2-Pentanone	8.573	43	168466	104.594	ug/l	98
52) Toluene	8.720	92	54946	19.590	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	29483	19.192	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	34477	21.022	ug/l	94
55) 1,1,2-Trichloroethane	9.152	97	21820	20.549	ug/l	98
56) Ethyl methacrylate	9.116	69	34572	20.540	ug/l	98
57) 1,3-Dichloropropane	9.311	76	38046	20.866	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	82255	95.211	ug/l	97
59) 2-Hexanone	9.433	43	126273	107.184	ug/l	97
60) Dibromochloromethane	9.524	129	22399	20.501	ug/l	97
61) 1,2-Dibromoethane	9.610	107	22958	20.464	ug/l	99
64) Tetrachloroethene	9.274	164	27291	19.825	ug/l	95
65) Chlorobenzene	10.079	112	56876	20.065	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.164	131	21416	20.538	ug/l	98
67) Ethyl Benzene	10.195	91	102848	20.500	ug/l	99
68) m/p-Xylenes	10.305	106	78672	40.757	ug/l	99
69) o-Xylene	10.640	106	37901	20.039	ug/l	99
70) Styrene	10.658	104	63938	20.843	ug/l	98
71) Bromoform	10.799	173	15465	20.297	ug/l #	99
73) Isopropylbenzene	10.963	105	100163	20.154	ug/l	99
74) N-amyl acetate	10.847	43	39958	20.122	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	28118	20.848	ug/l	99
76) 1,2,3-Trichloropropane	11.237	75	34186	24.751	ug/l	93
77) Bromobenzene	11.201	156	23731	20.087	ug/l	96
78) n-propylbenzene	11.305	91	112832	20.478	ug/l	99
79) 2-Chlorotoluene	11.366	91	67758	19.903	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	84077	20.554	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	7795	19.841	ug/l	98
82) 4-Chlorotoluene	11.457	91	78413	20.532	ug/l	98
83) tert-Butylbenzene	11.713	119	78564	20.152	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	83991	20.357	ug/l	98
85) sec-Butylbenzene	11.890	105	100435	20.323	ug/l	99
86) p-Isopropyltoluene	12.012	119	84899	20.279	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	43348	20.024	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	44696	20.473	ug/l	99
89) n-Butylbenzene	12.335	91	70580	20.605	ug/l	98
90) Hexachloroethane	12.542	117	12608	20.837	ug/l	96
91) 1,2-Dichlorobenzene	12.335	146	42882	20.530	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.944	75	6698	20.194	ug/l	91
93) 1,2,4-Trichlorobenzene	13.591	180	26776	20.312	ug/l	98
94) Hexachlorobutadiene	13.725	225	11454	20.542	ug/l	99
95) Naphthalene	13.780	128	95327	20.645	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	26874	20.133	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026815.D
Acq On : 08 Feb 2022 12:31
Operator : JC/MD
Sample : VX0208WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Feb 09 03:00:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 09 03:00:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:168 Resp: 100530
Ion Ratio Lower Upper
168 100
99 59.8 50.3 75.5



#2
Dichlorodifluoromethane
Concen: 19.820 ug/l
RT: 1.172 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31


Tgt Ion: 85 Resp: 21665
Ion Ratio Lower Upper
85 100
87 34.1 16.4 49.1








#3
Chloromethane
Concen: 18.690 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 50 Resp: 19006
Ion Ratio Lower Upper
50 100
52 32.8 26.1 39.1



#4
Vinyl Chloride
Concen: 18.957 ug/l
RT: 1.379 min Scan# 48
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 62 Resp: 20180
Ion Ratio Lower Upper
62 100
64 33.8 26.2 39.2









#5
Bromomethane
Concen: 21.261 ug/l
RT: 1.617 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 9604
Ion Ratio Lower Upper
94 100
96 100.1 75.9 113.9



#6
Chloroethane
Concen: 18.967 ug/l
RT: 1.690 min Scan# 99
Delta R.T. -0.001 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31


Tgt Ion: 64 Resp: 12109
Ion Ratio Lower Upper
64 100
66 29.6 25.4 38.2








#7
 Trichlorofluoromethane
 Concen: 19.998 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX026815.D
 Acq: 08 Feb 2022 12:31

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:101 Resp: 31327
 Ion Ratio Lower Upper
 101 100
 103 64.0 51.3 76.9



#8
 Diethyl Ether
 Concen: 19.976 ug/l
 RT: 2.135 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX026815.D
 Acq: 08 Feb 2022 12:31

Tgt Ion: 74 Resp: 11592
 Ion Ratio Lower Upper
 74 100
 45 98.6 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.526 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 21077
Ion Ratio Lower Upper
101 100
85 42.7 35.3 52.9
151 78.5 61.4 92.2



#10
Methyl Iodide
Concen: 17.463 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion:142 Resp: 22684
Ion Ratio Lower Upper
142 100
127 44.7 38.1 57.1
141 14.3 11.2 16.8





#11

Tert butyl alcohol

Concen: 100.130 ug/l

RT: 2.977 min Scan# 314

Delta R.T. -0.024 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 21829

Ion Ratio Lower Upper

59 100

57 10.7 7.2 10.8



#12

1,1-Dichloroethene

Concen: 19.206 ug/l

RT: 2.324 min Scan# 203

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion: 96 Resp: 19760


Ion Ratio Lower Upper

96 100

61 161.6 138.5 207.7

98 64.0 51.8 77.6





#13

Acrolein

Concen: 109.021 ug/l

RT: 2.239 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 21964

Ion Ratio Lower Upper

56 100

55 70.0 56.9 85.3



#14

Allyl chloride

Concen: 19.848 ug/l

RT: 2.672 min Scan# 260

Delta R.T. 0.006 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion: 41 Resp: 36172


Ion Ratio Lower Upper

41 100

39 64.2 54.2 81.4


76 31.9 25.2 37.8








#15
Acrylonitrile
Concen: 100.038 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 62496
Ion Ratio Lower Upper
53 100
52 82.5 66.6 100.0
51 35.7 29.5 44.3



#16
Acetone
Concen: 111.336 ug/l
RT: 2.385 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31


Tgt Ion: 43 Resp: 59010
Ion Ratio Lower Upper
43 100
58 31.3 23.0 34.6








#17
Carbon Disulfide
Concen: 19.204 ug/l
RT: 2.520 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 76 Resp: 55931
Ion Ratio Lower Upper
76 100
78 8.8 7.5 11.3



#18
Methyl Acetate
Concen: 19.061 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 43 Resp: 29068
Ion Ratio Lower Upper
43 100
74 23.1 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 19.952 ug/l
RT: 3.117 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 72659
Ion Ratio Lower Upper
73 100
57 21.7 17.4 26.2



#20
Methylene Chloride
Concen: 19.132 ug/l
RT: 2.794 min Scan# 280
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31


Tgt Ion: 84 Resp: 21972
Ion Ratio Lower Upper
84 100
49 121.2 101.5 152.3
51 37.8 31.2 46.8
86 65.7 51.3 76.9








#21
trans-1,2-Dichloroethene
Concen: 20.142 ug/l
RT: 3.099 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 21807
Ion Ratio Lower Upper
96 100
61 133.9 112.9 169.3
98 62.9 49.5 74.3




#22
Diisopropyl ether
Concen: 20.212 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31



Tgt Ion: 45 Resp: 74160
Ion Ratio Lower Upper
45 100
43 80.7 60.6 91.0
87 29.0 22.0 33.0
59 12.3 9.0 13.4









#23
 Vinyl Acetate
 Concen: 102.488 ug/l
 RT: 3.727 min Scan# 411
 Delta R.T. -0.000 min
 Lab File: VX026815.D
 Acq: 08 Feb 2022 12:31




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 316160
 Ion Ratio Lower Upper
 43 100
 86 10.7 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 19.850 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.006 min
 Lab File: VX026815.D
 Acq: 08 Feb 2022 12:31

Tgt Ion: 63 Resp: 39411
 Ion Ratio Lower Upper
 63 100
 98 7.4 3.7 11.1
 100 5.1 2.4 7.2









#25
2-Butanone
Concen: 104.492 ug/l
RT: 4.562 min Scan# 570
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 89231
Ion Ratio Lower Upper
43 100
72 26.0 20.4 30.6



#26
2,2-Dichloropropane
Concen: 21.646 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 77 Resp: 29037
Ion Ratio Lower Upper
77 100
97 23.6 11.0 33.0





#27

cis-1,2-Dichloroethene

Concen: 20.067 ug/l

RT: 4.495 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 24944

Ion Ratio Lower Upper

96 100

61 142.4 0.0 296.6

98 64.4 0.0 128.8



#28

Bromochloromethane

Concen: 20.576 ug/l

RT: 4.903 min Scan# 626

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion: 49 Resp: 16934


Ion Ratio Lower Upper

49 100

129 2.2 0.0 4.2

130 84.1 62.4 93.6





#29

Tetrahydrofuran

Concen: 99.447 ug/l

RT: 5.013 min Scan# 64

Delta R.T. 0.006 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 57733

Ion Ratio Lower Upper

42 100

72 44.9 34.6 52.0

71 41.5 31.8 47.8



#30

Chloroform

Concen: 19.653 ug/l

RT: 5.098 min Scan# 658

Delta R.T. -0.000 min

Lab File: VX026815.D




Acq: 08 Feb 2022 12:31


Tgt Ion: 83 Resp: 40540

Ion Ratio Lower Upper

83 100

85 66.1 51.6 77.4





#31

Cyclohexane

Concen: 20.261 ug/l

RT: 5.476 min Scan# 720

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 56 Resp: 38850

Ion Ratio Lower Upper

56 100

69 30.7 25.8 38.6

84 85.0 68.9 103.3



#32

1,1,1-Trichloroethane

Concen: 19.486 ug/l

RT: 5.391 min Scan# 706

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion: 97 Resp: 35439


Ion Ratio Lower Upper

97 100

99 64.2 51.3 76.9

61 42.7 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 51.092 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 61656

Ion Ratio Lower Upper

65 100

67 52.1 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion: 114 Resp: 161727


Ion Ratio Lower Upper

114 100

63 20.6 0.0 42.8


88 16.8 0.0 33.0








#35
Dibromofluoromethane
Concen: 51.546 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:113 Resp: 49666
Ion Ratio Lower Upper
113 100
111 102.2 81.2 121.8
192 19.4 15.0 22.4



#36
1,1-Dichloropropene
Concen: 20.103 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 75 Resp: 30558
Ion Ratio Lower Upper
75 100
110 36.1 17.2 51.6
77 31.5 24.6 36.8





#37

Ethyl Acetate

Concen: 19.529 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min


Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 43	Resp: 33736
Ion Ratio	Lower Upper
43	100
61	14.1 10.6 16.0
70	11.3 8.5 12.7



#38

Carbon Tetrachloride

Concen: 19.927 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Tgt Ion: 117	Resp: 31763
Ion Ratio	Lower Upper
117	100
119	92.9 74.7 112.1
121	31.5 25.4 38.0





#39

Methylcyclohexane

Concen: 20.655 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 38504

Ion Ratio Lower Upper

83 100

55 77.0 63.6 95.4

98 50.6 38.4 57.6



#40

Benzene

Concen: 19.944 ug/l

RT: 6.043 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX026815.D




Acq: 08 Feb 2022 12:31


Tgt Ion: 78 Resp: 88354

Ion Ratio Lower Upper

78 100


77 22.8 18.9 28.3








#41
Methacrylonitrile
Concen: 20.081 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41	Resp: 18801
Ion Ratio	Lower Upper
41 100	
39 57.0	45.4 68.0
67 71.4	53.2 79.8
52 30.1	23.0 34.6



#42
1,2-Dichloroethane
Concen: 20.850 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 62	Resp: 32549
Ion Ratio	Lower Upper
62 100	
98 9.9	0.0 17.8









#43
Isopropyl Acetate
Concen: 20.436 ug/l
RT: 6.348 min Scan# 863
Delta R.T. 0.006 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 54316
Ion Ratio Lower Upper
43 100
61 19.4 15.4 23.0
87 13.6 10.2 15.2



#44
Trichloroethene
Concen: 19.654 ug/l
RT: 7.128 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion:130 Resp: 26143
Ion Ratio Lower Upper
130 100
95 96.6 0.0 202.6





#45

1,2-Dichloropropane

Concen: 20.024 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 22444

Ion Ratio Lower Upper

63 100

65 32.0 23.8 35.6



#46

Dibromomethane

Concen: 20.200 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion: 93 Resp: 15667


Ion Ratio Lower Upper

93 100

95 82.4 66.1 99.1


174 104.7 78.7 118.1








#47
Bromodichloromethane
Concen: 20.662 ug/l
RT: 7.823 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 83 Resp: 30688
Ion Ratio Lower Upper
83 100
85 66.4 51.4 77.0
127 9.2 6.9 10.3



#48
Methyl methacrylate
Concen: 20.276 ug/l
RT: 7.695 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 41 Resp: 26703
Ion Ratio Lower Upper
41 100
69 79.8 62.2 93.4
39 54.8 45.4 68.0





#49

1,4-Dioxane

Concen: 452.278 ug/l

RT: 7.677 min Scan# 1081

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 10264

Ion Ratio Lower Upper

88 100

43 32.5 29.9 44.9

58 68.2 57.8 86.6



#50

Toluene-d8

Concen: 50.922 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX026815.D




Acq: 08 Feb 2022 12:31


Tgt Ion: 98 Resp: 180105

Ion Ratio Lower Upper

98 100

100 66.8 53.2 79.8









#51
4-Methyl-2-Pentanone
Concen: 104.594 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 168466
Ion Ratio Lower Upper
43 100
58 39.3 30.5 45.7



#52
Toluene
Concen: 19.590 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 92 Resp: 54946
Ion Ratio Lower Upper
92 100
91 169.7 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 19.192 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 29483

Ion Ratio Lower Upper

75 100

77 31.1 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 21.022 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion: 75 Resp: 34477


Ion Ratio Lower Upper

75 100

77 31.2 26.2 39.2

39 49.2 43.5 65.3





#55

1,1,2-Trichloroethane

Concen: 20.549 ug/l

RT: 9.152 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 97 Resp: 21820





Ion Ratio Lower Upper

97 100

83 82.0 68.4 102.6

85 54.7 44.0 66.0

99 63.1 49.8 74.6



#56

Ethyl methacrylate

Concen: 20.540 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion: 69 Resp: 34572


Ion Ratio Lower Upper

69 100

41 69.9 57.2 85.8

39 36.8 30.9 46.3





#57

1,3-Dichloropropane

Concen: 20.866 ug/l

RT: 9.311 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 38046

Ion Ratio Lower Upper

76 100

78 32.3 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 95.211 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX026815.D




Acq: 08 Feb 2022 12:31


Tgt Ion: 63 Resp: 82255

Ion Ratio Lower Upper

63 100

106 30.2 23.0 34.6





#59

2-Hexanone

Concen: 107.184 ug/l

RT: 9.433 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 126273

Ion Ratio Lower Upper

43 100

58 55.1 26.4 79.2



#60

Dibromochloromethane

Concen: 20.501 ug/l

RT: 9.524 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX026815.D




Acq: 08 Feb 2022 12:31


Tgt Ion: 129 Resp: 22399

Ion Ratio Lower Upper

129 100

127 79.4 38.5 115.5









#61
1,2-Dibromoethane
Concen: 20.464 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 22958
Ion Ratio Lower Upper
107 100
109 95.3 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 52.540 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31


Tgt Ion: 95 Resp: 65424
Ion Ratio Lower Upper
95 100
174 75.8 0.0 148.6
176 73.1 0.0 141.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 137064
Ion Ratio Lower Upper
117 100
82 53.8 45.2 67.8
119 31.8 26.2 39.2




#64
Tetrachloroethene
Concen: 19.825 ug/l
RT: 9.274 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31



Tgt Ion:164 Resp: 27291
Ion Ratio Lower Upper
164 100
166 126.3 101.8 152.6
129 95.5 83.3 124.9
131 91.3 78.0 117.0









#65
Chlorobenzene
Concen: 20.065 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 56876
Ion Ratio Lower Upper
112 100
114 31.4 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 20.538 ug/l
RT: 10.164 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion:131 Resp: 21416
Ion Ratio Lower Upper
131 100
133 94.0 47.9 143.8
119 64.1 31.7 95.1









#67
Ethyl Benzene
Concen: 20.500 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 102848
Ion Ratio Lower Upper
91 100
106 31.5 24.7 37.1



#68
m/p-Xylenes
Concen: 40.757 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 106 Resp: 78672
Ion Ratio Lower Upper
106 100
91 204.7 165.3 247.9









#69
o-Xylene
Concen: 20.039 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 37901
Ion Ratio Lower Upper
106 100
91 218.5 109.7 329.0



#70
Styrene
Concen: 20.843 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion:104 Resp: 63938
Ion Ratio Lower Upper
104 100
78 50.7 42.5 63.7
103 54.7 43.3 64.9









#71
Bromoform
Concen: 20.297 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:173 Resp: 15465
Ion Ratio Lower Upper
173 100
175 49.0 24.3 72.9
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion:152 Resp: 63097
Ion Ratio Lower Upper
152 100
115 70.0 44.2 132.6
150 166.6 0.0 348.2









#73
Isopropylbenzene
Concen: 20.154 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 100163
Ion Ratio Lower Upper
105 100
120 26.8 13.1 39.1



#74
N-ethyl acetate
Concen: 20.122 ug/l
RT: 10.847 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 43 Resp: 39958
Ion Ratio Lower Upper
43 100
70 42.5 32.9 49.3
55 24.1 18.3 27.5
61 24.1 18.9 28.3









#75
 1,1,2,2-Tetrachloroethane
 Concen: 20.848 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. -0.000 min
 Lab File: VX026815.D
 Acq: 08 Feb 2022 12:31




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 83 Resp: 28118
 Ion Ratio Lower Upper
 83 100
 131 10.3 5.3 15.9
 85 65.1 32.0 96.2



#76
 1,2,3-Trichloropropane
 Concen: 24.751 ug/l
 RT: 11.237 min Scan# 1665
 Delta R.T. -0.006 min
 Lab File: VX026815.D
 Acq: 08 Feb 2022 12:31

Tgt Ion: 75 Resp: 34186
 Ion Ratio Lower Upper
 75 100
 77 32.3 18.3 54.8









#77
Bromobenzene
Concen: 20.087 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 23731
Ion Ratio Lower Upper
156 100
77 160.8 84.5 253.4
158 96.5 48.1 144.4



#78
n-propylbenzene
Concen: 20.478 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 91 Resp: 112832
Ion Ratio Lower Upper
91 100
120 23.5 11.6 34.7









#79
2-Chlorotoluene
Concen: 19.903 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 67758
Ion Ratio Lower Upper
91 100
126 34.0 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 20.554 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 105 Resp: 84077
Ion Ratio Lower Upper
105 100
120 48.7 24.0 72.0









#81
trans-1,4-Dichloro-2-butene
Concen: 19.841 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 7795
Ion Ratio Lower Upper
75 100
53 118.2 94.0 141.0
89 50.0 37.0 55.4



#82
4-Chlorotoluene
Concen: 20.532 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 91 Resp: 78413
Ion Ratio Lower Upper
91 100
126 29.5 14.1 42.4









#83
tert-Butylbenzene
Concen: 20.152 ug/l
RT: 11.713 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 78564
Ion Ratio Lower Upper
119 100
91 60.7 31.4 94.0
134 23.9 11.8 35.3



#84
1,2,4-Trimethylbenzene
Concen: 20.357 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31


Tgt Ion:105 Resp: 83991
Ion Ratio Lower Upper
105 100
120 44.8 21.9 65.5








#85
 sec-Butylbenzene
 Concen: 20.323 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX026815.D
 Acq: 08 Feb 2022 12:31

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:105 Resp: 100435
 Ion Ratio Lower Upper
 105 100
 134 20.4 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 20.279 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX026815.D
 Acq: 08 Feb 2022 12:31

Tgt Ion:119 Resp: 84899
 Ion Ratio Lower Upper
 119 100
 134 26.3 13.0 39.0
 91 24.7 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 20.024 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 43348

Ion Ratio Lower Upper

146 100

111 41.6 21.3 63.7

148 63.7 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 20.473 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion:146 Resp: 44696


Ion Ratio Lower Upper

146 100

111 40.5 20.2 60.6


148 63.6 31.4 94.0








#89
n-Butylbenzene
Concen: 20.605 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 91 Resp: 70580
Ion Ratio Lower Upper
91 100
92 55.5 27.2 81.6
134 27.1 13.1 39.1



#90
Hexachloroethane
Concen: 20.837 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion: 117 Resp: 12608
Ion Ratio Lower Upper
117 100
201 74.6 35.6 106.8





#91

1,2-Dichlorobenzene

Concen: 20.530 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 42882

Ion Ratio Lower Upper

146 100

111 41.5 21.8 65.3

148 64.1 32.0 96.2



#92

1,2-Dibromo-3-Chloropropane

Concen: 20.194 ug/l

RT: 12.944 min Scan# 1945

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion: 75 Resp: 6698


Ion Ratio Lower Upper

75 100

155 89.2 40.0 119.9

157 111.1 51.6 154.8





#93

1,2,4-Trichlorobenzene

Concen: 20.312 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. -0.000 min


Lab File: VX026815.D

Acq: 08 Feb 2022 12:31

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:180 Resp: 26776

Ion Ratio Lower Upper

180 100

182 95.7 46.7 140.0

145 32.5 16.3 48.8



#94

Hexachlorobutadiene

Concen: 20.542 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX026815.D

Acq: 08 Feb 2022 12:31




Tgt Ion:225 Resp: 11454


Ion Ratio Lower Upper

225 100

223 63.7 32.0 96.0


227 66.2 32.3 96.9








#95
Naphthalene
Concen: 20.645 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 95327
Ion Ratio Lower Upper
128 100
127 12.7 10.3 15.5
129 11.0 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 20.133 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX026815.D
Acq: 08 Feb 2022 12:31

Tgt Ion:180 Resp: 26874
Ion Ratio Lower Upper
180 100
182 95.6 47.2 141.6
145 34.6 17.5 52.5

