

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026808.D
 Acq On : 08 Feb 2022 02:38
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Feb 08 07:18:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218027	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192526	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84370	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80048	51.007	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.020%		
35) Dibromofluoromethane	5.391	113	66771	51.403	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.800%		
50) Toluene-d8	8.653	98	248570	52.132	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 104.260%		
62) 4-Bromofluorobenzene	11.079	95	91011	54.215	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 108.440%		
Target Compounds						
					Qvalue	
10) Methyl Iodide	2.465	142	828	2.681	ug/l	90
16) Acetone	2.392	43	708	1.027	ug/l #	89
17) Carbon Disulfide	2.520	76	4224	1.115	ug/l	95
53) t-1,3-Dichloropropene	8.982	75	305	1.487	ug/l	99
94) Hexachlorobutadiene	13.725	225	1177	1.579	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026808.D
Acq On : 08 Feb 2022 02:38
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Feb 08 07:18:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

