

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026816.D
 Acq On : 08 Feb 2022 12:55
 Operator : JC/MD
 Sample : N1368-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31382

Quant Time: Feb 09 03:01:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	220362	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191725	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85248	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81419	50.767	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.540%
35) Dibromofluoromethane	5.391	113	68729	52.350	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.700%
50) Toluene-d8	8.653	98	253538	52.610	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.220%
62) 4-Bromofluorobenzene	11.079	95	90884	53.566	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.140%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	99808	141.695	ug/l	97
17) Carbon Disulfide	2.514	76	1042	0.269	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	1047	0.216	ug/l	97
20) Methylene Chloride	2.800	84	404	0.265	ug/l #	82
25) 2-Butanone	4.562	43	24939	21.975	ug/l	99
40) Benzene	6.043	78	24173	4.005	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	20633	9.402	ug/l	98
52) Toluene	8.720	92	169841	44.441	ug/l	99
67) Ethyl Benzene	10.195	91	63282	9.017	ug/l	98
68) m/p-Xylenes	10.299	106	107110	39.670	ug/l	97
69) o-Xylene	10.640	106	57554	21.754	ug/l	98
73) Isopropylbenzene	10.963	105	12688	1.890	ug/l	99
78) n-propylbenzene	11.305	91	43911	5.899	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	75994	13.751	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	323951	58.115	ug/l	97
85) sec-Butylbenzene	11.890	105	19032	2.850	ug/l	87
86) p-Isopropyltoluene	12.012	119	13908	2.459	ug/l	99
89) n-Butylbenzene	12.329	91	41550	8.978	ug/l #	40
95) Naphthalene	13.780	128	82847	13.280	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026816.D
Acq On : 08 Feb 2022 12:55
Operator : JC/MD
Sample : N1368-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31382

Quant Time: Feb 09 03:01:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

