

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026820.D
 Acq On : 08 Feb 2022 14:27
 Operator : JC/MD
 Sample : N1411-02
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 282-284

Quant Time: Feb 09 03:03:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121879	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198315	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	174445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	73192	50.028	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.060%	
35) Dibromofluoromethane	5.397	113	59924	50.718	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.440%	
50) Toluene-d8	8.653	98	228675	52.726	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.460%	
62) 4-Bromofluorobenzene	11.085	95	85062	55.708	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.420%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026820.D
Acq On : 08 Feb 2022 14:27
Operator : JC/MD
Sample : N1411-02
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
282-284

Quant Time: Feb 09 03:03:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

