

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX020822\  
 Data File : VX026823.D  
 Acq On : 08 Feb 2022 15:37  
 Operator : JC/MD  
 Sample : N1398-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Feb 09 03:04:03 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X020822W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Feb 08 07:07:46 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min)  |
|-----------------------------|--------|-------|----------|----------|------------|-----------|
| -----                       |        |       |          |          |            |           |
| Internal Standards          |        |       |          |          |            |           |
| 1) Pentafluorobenzene       | 5.562  | 168   | 128210   | 50.000   | ug/l       | 0.00      |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 213147   | 50.000   | ug/l       | 0.00      |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 185944   | 50.000   | ug/l       | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 80404    | 50.000   | ug/l       | 0.00      |
| System Monitoring Compounds |        |       |          |          |            |           |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 79393    | 51.586   | ug/l       | 0.00      |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 103.180% |           |
| 35) Dibromofluoromethane    | 5.391  | 113   | 65691    | 51.730   | ug/l       | 0.00      |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 103.460% |           |
| 50) Toluene-d8              | 8.653  | 98    | 242922   | 52.114   | ug/l       | 0.00      |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 104.220% |           |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 87487    | 53.309   | ug/l       | 0.00      |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 106.620% |           |
| Target Compounds            |        |       |          |          |            |           |
| 16) Acetone                 | 2.385  | 43    | 1937     | 2.866    | ug/l       | Qvalue 96 |
| -----                       |        |       |          |          |            |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX020822\  
Data File : VX026823.D  
Acq On : 08 Feb 2022 15:37  
Operator : JC/MD  
Sample : N1398-01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
STORAGE-BLANK-SOIL-REF-6

Quant Time: Feb 09 03:04:03 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X020822W.M  
Quant Title : SW846 8260  
QLast Update : Tue Feb 08 07:07:46 2022  
Response via : Initial Calibration

