

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026824.D
 Acq On : 08 Feb 2022 16:00
 Operator : JC/MD
 Sample : N1398-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 09 03:04:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	140144	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	232961	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85447	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87890	52.244	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.480%
35) Dibromofluoromethane	5.391	113	72470	52.214	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.420%
50) Toluene-d8	8.653	98	266254	52.261	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.520%
62) 4-Bromofluorobenzene	11.079	95	94036	52.426	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.860%
Target Compounds						
16) Acetone	2.386	43	1977	2.676	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026824.D
Acq On : 08 Feb 2022 16:00
Operator : JC/MD
Sample : N1398-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Feb 09 03:04:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

