

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026825.D
 Acq On : 08 Feb 2022 16:23
 Operator : JC/MD
 Sample : N1398-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 09 03:04:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	133874	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	225319	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	85938	53.477	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 106.960%		
35) Dibromofluoromethane	5.391	113	70598	52.591	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.180%		
50) Toluene-d8	8.653	98	256185	51.990	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 103.980%		
62) 4-Bromofluorobenzene	11.079	95	94697	54.585	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 109.180%		
Target Compounds						Qvalue
16) Acetone	2.386	43	1790	2.536	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026825.D
Acq On : 08 Feb 2022 16:23
Operator : JC/MD
Sample : N1398-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Feb 09 03:04:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

