

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026827.D
 Acq On : 08 Feb 2022 17:10
 Operator : JC/MD
 Sample : N1400-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Feb 09 03:05:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	126508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	211646	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183175	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80552	53.044	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.080%	
35) Dibromofluoromethane	5.391	113	65541	51.978	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.960%	
50) Toluene-d8	8.653	98	241240	52.120	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.240%	
62) 4-Bromofluorobenzene	11.079	95	86944	53.354	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.700%	
Target Compounds						
16) Acetone	2.386	43	1986	2.978	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026827.D
Acq On : 08 Feb 2022 17:10
Operator : JC/MD
Sample : N1400-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Feb 09 03:05:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

