

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026828.D
 Acq On : 08 Feb 2022 17:33
 Operator : JC/MD
 Sample : N1429-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-21S

Quant Time: Feb 09 03:05:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127077	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	210543	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185895	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80049	52.476	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.960%	
35) Dibromofluoromethane	5.391	113	65915	52.548	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.100%	
50) Toluene-d8	8.653	98	244496	53.100	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.200%	
62) 4-Bromofluorobenzene	11.079	95	88222	54.422	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.840%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	11260	8.368	ug/l	100
16) Acetone	2.386	43	2195	3.276	ug/l	94
21) trans-1,2-Dichloroethene	3.099	96	1030	0.753	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	208436	132.652	ug/l	87
44) Trichloroethene	7.129	130	29895	17.264	ug/l	99
64) Tetrachloroethene	9.275	164	46591	24.954	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026828.D
Acq On : 08 Feb 2022 17:33
Operator : JC/MD
Sample : N1429-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-21S

Quant Time: Feb 09 03:05:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

