

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026829.D
 Acq On : 08 Feb 2022 17:56
 Operator : JC/MD
 Sample : N1429-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-21D

Quant Time: Feb 09 03:05:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120425	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	201298	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176772	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79926	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	77280	53.459	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.920%	
35) Dibromofluoromethane	5.391	113	64096	53.445	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.880%	
50) Toluene-d8	8.653	98	231577	52.604	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.200%	
62) 4-Bromofluorobenzene	11.079	95	86140	55.578	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.160%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	98845	77.514	ug/l	99
12) 1,1-Dichloroethene	2.325	96	1067	0.866	ug/l #	91
16) Acetone	2.380	43	3589	5.653	ug/l	99
21) trans-1,2-Dichloroethene	3.099	96	10155	7.830	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	895942	601.688	ug/l	88
39) Methylcyclohexane	7.379	83	1704	0.734	ug/l #	58
40) Benzene	6.044	78	4512	0.818	ug/l	98
44) Trichloroethene	7.135	130	1519	0.917	ug/l	88
64) Tetrachloroethene	9.275	164	804	0.453	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026829.D
Acq On : 08 Feb 2022 17:56
Operator : JC/MD
Sample : N1429-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-21D

Quant Time: Feb 09 03:05:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

