

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026834.D
 Acq On : 08 Feb 2022 19:52
 Operator : JC/MD
 Sample : N1400-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KTS1-WC-L-MH-02

Quant Time: Feb 09 03:06:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	140852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	235919	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	89897	53.169	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.340%	
35) Dibromofluoromethane	5.391	113	74516	53.015	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.040%	
50) Toluene-d8	8.653	98	264940	51.351	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.700%	
62) 4-Bromofluorobenzene	11.079	95	95347	52.491	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.980%	
Target Compounds						Qvalue
16) Acetone	2.386	43	6268	8.441	ug/l	95
17) Carbon Disulfide	2.520	76	4455	1.092	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026834.D
Acq On : 08 Feb 2022 19:52
Operator : JC/MD
Sample : N1400-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KTS1-WC-L-MH-02

Quant Time: Feb 09 03:06:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

