

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026835.D
 Acq On : 08 Feb 2022 20:15
 Operator : JC/MD
 Sample : N1434-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 34040

Quant Time: Feb 09 03:07:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	123045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	203868	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	78810	53.357	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.720%	
35) Dibromofluoromethane	5.391	113	64188	52.847	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.700%	
50) Toluene-d8	8.653	98	233430	52.357	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.720%	
62) 4-Bromofluorobenzene	11.079	95	86222	54.930	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.860%	
Target Compounds						Qvalue
16) Acetone	2.386	43	850	1.310	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026835.D
Acq On : 08 Feb 2022 20:15
Operator : JC/MD
Sample : N1434-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
34040

Quant Time: Feb 09 03:07:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

