

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026840.D
 Acq On : 08 Feb 2022 22:57
 Operator : JC/MD
 Sample : VX0208WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0208WBL02

Quant Time: Feb 09 03:09:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	129692	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	218222	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83737	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81791	52.537	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.080%
35) Dibromofluoromethane	5.391	113	67494	51.914	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.820%
50) Toluene-d8	8.653	98	252612	52.932	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.860%
62) 4-Bromofluorobenzene	11.079	95	92372	54.977	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026840.D
Acq On : 08 Feb 2022 22:57
Operator : JC/MD
Sample : VX0208WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0208WBL02

Quant Time: Feb 09 03:09:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

