

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026844.D
 Acq On : 09 Feb 2022 00:29
 Operator : JC/MD
 Sample : N1425-05 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20722

Quant Time: Feb 09 05:19:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222291	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84281	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84697	53.254	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.500%
35) Dibromofluoromethane	5.391	113	68534	51.749	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.500%
50) Toluene-d8	8.653	98	256617	52.787	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.580%
62) 4-Bromofluorobenzene	11.079	95	92916	54.288	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026844.D
Acq On : 09 Feb 2022 00:29
Operator : JC/MD
Sample : N1425-05 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20722

Quant Time: Feb 09 05:19:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

