

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026849.D
 Acq On : 09 Feb 2022 02:25
 Operator : JC/MD
 Sample : PB142539ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142539ZHE#02

Quant Time: Feb 09 05:20:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	135396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227693	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202436	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87241	53.677	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.360%	
35) Dibromofluoromethane	5.391	113	71331	52.583	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.160%	
50) Toluene-d8	8.653	98	265629	53.344	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.680%	
62) 4-Bromofluorobenzene	11.079	95	95950	54.731	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.460%	
Target Compounds						Qvalue
16) Acetone	2.386	43	4995	6.997	ug/l	93
20) Methylene Chloride	2.794	84	1599	1.034	ug/l	87
43) Isopropyl Acetate	6.348	43	12355	3.302	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026849.D
Acq On : 09 Feb 2022 02:25
Operator : JC/MD
Sample : PB142539ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142539ZHE#02

Quant Time: Feb 09 05:20:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

