

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026853.D
 Acq On : 09 Feb 2022 03:57
 Operator : JC/MD
 Sample : PB142539ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142539ZHE#06

Quant Time: Feb 09 05:21:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	223512	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202977	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85332	53.514	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.020%	
35) Dibromofluoromethane	5.391	113	68916	51.753	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.500%	
50) Toluene-d8	8.653	98	264232	54.056	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.120%	
62) 4-Bromofluorobenzene	11.079	95	97160	56.458	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.920%	
Target Compounds						Qvalue
16) Acetone	2.392	43	5064	7.231	ug/l	96
43) Isopropyl Acetate	6.342	43	12233	3.330	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026853.D
Acq On : 09 Feb 2022 03:57
Operator : JC/MD
Sample : PB142539ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142539ZHE#06

Quant Time: Feb 09 05:21:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

