

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026854.D
 Acq On : 09 Feb 2022 04:20
 Operator : JC/MD
 Sample : PB142539ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142539ZHE#07

Quant Time: Feb 09 05:22:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129884	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218769	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87847	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84417	54.144	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 108.280%			
35) Dibromofluoromethane	5.391	113	67642	51.897	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.800%			
50) Toluene-d8	8.653	98	258474	54.025	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 108.040%			
62) 4-Bromofluorobenzene	11.079	95	97385	57.816	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 115.640%			
Target Compounds						
16) Acetone	2.386	43	5141	7.508	ug/l	99
20) Methylene Chloride	2.788	84	1515	1.021	ug/l #	91
43) Isopropyl Acetate	6.342	43	12088	3.362	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026854.D
Acq On : 09 Feb 2022 04:20
Operator : JC/MD
Sample : PB142539ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142539ZHE#07

Quant Time: Feb 09 05:22:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

