

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026856.D
 Acq On : 09 Feb 2022 05:06
 Operator : JC/MD
 Sample : N1432-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-2

Quant Time: Feb 09 06:43:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132141	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221069	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87662	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85376	53.824	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.640%	
35) Dibromofluoromethane	5.391	113	68821	52.253	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.500%	
50) Toluene-d8	8.653	98	261277	54.043	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.080%	
62) 4-Bromofluorobenzene	11.079	95	97691	57.394	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 114.780%	
Target Compounds						Qvalue
16) Acetone	2.386	43	4994	7.168	ug/l	100
20) Methylene Chloride	2.794	84	1786	1.183	ug/l	93
43) Isopropyl Acetate	6.342	43	11416	3.142	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026856.D
Acq On : 09 Feb 2022 05:06
Operator : JC/MD
Sample : N1432-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-2

Quant Time: Feb 09 06:43:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

