

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX020822\
 Data File : VX026837.D
 Acq On : 08 Feb 2022 21:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 09 03:07:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	50.000	50.215	-0.4	100	0.00
3 P	Chloromethane	50.000	48.223	3.6	99	0.00
4 C	Vinyl Chloride	50.000	49.371	1.3#	99	0.00
5 T	Bromomethane	50.000	43.333	13.3	94	0.00
6 T	Chloroethane	50.000	54.014	-8.0	100	0.00
7 T	Trichlorofluoromethane	50.000	48.780	2.4	97	0.00
8 T	Diethyl Ether	50.000	49.513	1.0	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.308	1.4	99	0.00
10 T	Methyl Iodide	50.000	43.070	13.9	89	0.00
11 T	Tert butyl alcohol	250.000	248.889	0.4	103	-0.02
12 CM	1,1-Dichloroethene	50.000	49.021	2.0#	99	0.00
13 T	Acrolein	250.000	259.843	-3.9	105	0.00
14 T	Allyl chloride	50.000	50.344	-0.7	101	0.00
15 T	Acrylonitrile	250.000	243.285	2.7	98	0.00
16 T	Acetone	250.000	233.830	6.5	99	0.00
17 T	Carbon Disulfide	50.000	48.280	3.4	99	0.00
18 T	Methyl Acetate	50.000	46.342	7.3	98	0.00
19 T	Methyl tert-butyl Ether	50.000	49.255	1.5	101	0.00
20 T	Methylene Chloride	50.000	48.572	2.9	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.200	-0.4	100	0.00
22 T	Diisopropyl ether	50.000	49.666	0.7	100	0.00
23 T	Vinyl Acetate	250.000	257.417	-3.0	101	0.00
24 P	1,1-Dichloroethane	50.000	50.220	-0.4	101	0.00
25 T	2-Butanone	250.000	246.261	1.5	99	0.00
26 T	2,2-Dichloropropane	50.000	53.866	-7.7	105	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.401	-0.8	100	0.00
28 T	Bromochloromethane	50.000	49.422	1.2	100	0.00
29 T	Tetrahydrofuran	250.000	242.457	3.0	99	0.00
30 C	Chloroform	50.000	49.349	1.3#	102	0.00
31 T	Cyclohexane	50.000	49.424	1.2	100	0.00
32 T	1,1,1-Trichloroethane	50.000	50.265	-0.5	101	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.143	1.7	105	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	49.407	1.2	105	0.00
36 T	1,1-Dichloropropene	50.000	49.635	0.7	100	0.00
37 T	Ethyl Acetate	50.000	48.729	2.5	99	0.00
38 T	Carbon Tetrachloride	50.000	50.375	-0.8	101	0.00
39 T	Methylcyclohexane	50.000	50.844	-1.7	101	0.00
40 TM	Benzene	50.000	49.799	0.4	100	0.00
41 T	Methacrylonitrile	50.000	49.816	0.4	99	0.00
42 TM	1,2-Dichloroethane	50.000	51.378	-2.8	102	0.00
43 T	Isopropyl Acetate	50.000	51.259	-2.5	100	0.00
44 TM	Trichloroethene	50.000	49.035	1.9	98	0.00
45 C	1,2-Dichloropropane	50.000	50.175	-0.3#	101	0.00
46 T	Dibromomethane	50.000	50.939	-1.9	102	0.00
47 T	Bromodichloromethane	50.000	52.268	-4.5	102	0.00
48 T	Methyl methacrylate	50.000	49.883	0.2	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
 Data File : VX026837.D
 Acq On : 08 Feb 2022 21:01
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 09 03:07:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1036.094	-3.6	103	0.00
50 S	Toluene-d8	50.000	49.609	0.8	105	0.00
51 T	4-Methyl-2-Pentanone	250.000	249.757	0.1	101	0.00
52 CM	Toluene	50.000	49.410	1.2#	101	0.00
53 T	t-1,3-Dichloropropene	50.000	49.177	1.6	103	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.699	-7.4	101	0.00
55 T	1,1,2-Trichloroethane	50.000	51.237	-2.5	102	0.00
56 T	Ethyl methacrylate	50.000	51.988	-4.0	102	0.00
57 T	1,3-Dichloropropane	50.000	51.158	-2.3	101	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	257.303	-2.9	95	0.00
59 T	2-Hexanone	250.000	251.676	-0.7	101	0.00
60 T	Dibromochloromethane	50.000	53.121	-6.2	102	0.00
61 T	1,2-Dibromoethane	50.000	51.566	-3.1	102	0.00
62 S	4-Bromofluorobenzene	50.000	49.037	1.9	104	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	98	0.00
64 T	Tetrachloroethene	50.000	49.609	0.8	96	0.00
65 PM	Chlorobenzene	50.000	50.089	-0.2	102	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.286	-2.6	102	0.00
67 C	Ethyl Benzene	50.000	51.073	-2.1#	102	0.00
68 T	m/p-Xylenes	100.000	100.889	-0.9	102	0.00
69 T	o-Xylene	50.000	50.012	-0.0	101	0.00
70 T	Styrene	50.000	50.515	-1.0	100	0.00
71 P	Bromoform	50.000	47.753	4.5	103	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	50.930	-1.9	101	0.00
74 T	N-ethyl acetate	50.000	52.091	-4.2	102	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.814	0.4	105	0.00
76 T	1,2,3-Trichloropropane	50.000	49.665	0.7	100	0.00
77 T	Bromobenzene	50.000	49.763	0.5	101	0.00
78 T	n-propylbenzene	50.000	51.675	-3.3	102	0.00
79 T	2-Chlorotoluene	50.000	49.776	0.4	102	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.604	-1.2	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.140	3.7	101	0.00
82 T	4-Chlorotoluene	50.000	50.401	-0.8	101	0.00
83 T	tert-Butylbenzene	50.000	50.844	-1.7	103	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.232	-0.5	102	0.00
85 T	sec-Butylbenzene	50.000	51.013	-2.0	102	0.00
86 T	p-Isopropyltoluene	50.000	50.786	-1.6	101	0.00
87 T	1,3-Dichlorobenzene	50.000	49.868	0.3	102	0.00
88 T	1,4-Dichlorobenzene	50.000	49.385	1.2	101	0.00
89 T	n-Butylbenzene	50.000	51.972	-3.9	102	0.00
90 T	Hexachloroethane	50.000	53.157	-6.3	103	0.00
91 T	1,2-Dichlorobenzene	50.000	50.426	-0.9	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.565	-3.1	99	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.419	-2.8	101	0.00
94 T	Hexachlorobutadiene	50.000	48.519	3.0	97	0.00
95 T	Naphthalene	50.000	51.828	-3.7	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020822\
Data File : VX026837.D
Acq On : 08 Feb 2022 21:01
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Feb 09 03:07:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.789	-1.6	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6