

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020823\
 Data File : VX034057.D
 Acq On : 07 Feb 2023 13:24
 Operator : JC/MD
 Sample : 01394-04ME
 Misc : 3.65g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C1152ME

Quant Time: Feb 07 20:43:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML020223WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Feb 07 20:42:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	210925	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	176501	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	98747	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	52865	37.633	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.260%
7) Chloroethane-d5	1.648	69	24940	29.245	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	58.480%#
11) 1,1-Dichloroethene-d2	2.288	63	79333	30.821	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.640%
21) 2-Butanone-d5	4.465	46	86844	107.829	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.830%
24) Chloroform-d	5.056	84	117957	46.812	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.620%
26) 1,2-Dichloroethane-d4	5.952	65	80123	52.700	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.400%
32) Benzene-d6	5.970	84	248819	49.179	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.360%
36) 1,2-Dichloropropane-d6	7.306	67	77117	51.556	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.120%
41) Toluene-d8	8.647	98	229195	47.754	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.500%
43) trans-1,3-Dichloroprop...	8.952	79	30640	45.138	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.280%
47) 2-Hexanone-d5	9.403	63	63320	112.885	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.880%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	105904	56.463	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	112.920%
66) 1,2-Dichlorobenzene-d4	12.323	152	106736	54.003	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020823\
Data File : VX034057.D
Acq On : 07 Feb 2023 13:24
Operator : JC/MD
Sample : 01394-04ME
Misc : 3.65g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C1152ME

Quant Time: Feb 07 20:43:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML020223WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Feb 07 20:42:13 2023
Response via : Initial Calibration

