

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026888.D
 Acq On : 09 Feb 2022 23:52
 Operator : JC/MD
 Sample : VX0209WBS02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 10 01:13:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	100419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	168293	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	151219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66798	55.414	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.820%	
35) Dibromofluoromethane	5.391	113	52506	52.367	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.740%	
50) Toluene-d8	8.653	98	200221	54.401	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.800%	
62) 4-Bromofluorobenzene	11.079	95	72163	55.691	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.380%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	21463	19.656	ug/l	100
3) Chloromethane	1.294	50	21007	20.681	ug/l	98
4) Vinyl Chloride	1.380	62	21720	20.426	ug/l	99
5) Bromomethane	1.617	94	10126	22.442	ug/l	100
6) Chloroethane	1.690	64	12304	19.316	ug/l	98
7) Trichlorofluoromethane	1.892	101	30589	19.548	ug/l	98
8) Diethyl Ether	2.135	74	11222	19.360	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.337	101	19628	19.136	ug/l	99
10) Methyl Iodide	2.459	142	23935	18.320	ug/l	98
11) Tert butyl alcohol	2.977	59	19229	88.302	ug/l	97
12) 1,1-Dichloroethene	2.324	96	19413	18.890	ug/l	94
13) Acrolein	2.239	56	17438	86.651	ug/l	98
14) Allyl chloride	2.666	41	35158	19.313	ug/l	95
15) Acrylonitrile	3.068	53	63429	101.644	ug/l	98
16) Acetone	2.385	43	53194	100.473	ug/l	96
17) Carbon Disulfide	2.520	76	51914	17.844	ug/l	100
18) Methyl Acetate	2.709	43	28708	18.846	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	68449	18.817	ug/l	97
20) Methylene Chloride	2.794	84	22111	19.274	ug/l	97
21) trans-1,2-Dichloroethene	3.099	96	20831	19.262	ug/l	97
22) Diisopropyl ether	3.763	45	71224	19.433	ug/l	96
23) Vinyl Acetate	3.727	43	302214	98.075	ug/l	99
24) 1,1-Dichloroethane	3.617	63	37945	19.133	ug/l	99
25) 2-Butanone	4.562	43	87530	102.614	ug/l	100
26) 2,2-Dichloropropane	4.483	77	25384	18.944	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	24118	19.424	ug/l	96
28) Bromochloromethane	4.903	49	16940	20.606	ug/l	100
29) Tetrahydrofuran	5.007	42	58864	101.507	ug/l	99
30) Chloroform	5.098	83	39887	19.358	ug/l	100
31) Cyclohexane	5.476	56	36175	18.887	ug/l	94
32) 1,1,1-Trichloroethane	5.391	97	33953	18.690	ug/l	99
36) 1,1-Dichloropropene	5.696	75	29515	18.660	ug/l	99
37) Ethyl Acetate	4.714	43	33992	18.910	ug/l	99
38) Carbon Tetrachloride	5.684	117	30213	18.215	ug/l	95
39) Methylcyclohexane	7.385	83	35800	18.456	ug/l	98
40) Benzene	6.043	78	85489	18.544	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026888.D
 Acq On : 09 Feb 2022 23:52
 Operator : JC/MD
 Sample : VX0209WBS02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 10 01:13:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	18663	19.156	ug/l	98
42) 1,2-Dichloroethane	6.092	62	31999	19.698	ug/l	98
43) Isopropyl Acetate	6.342	43	52082	18.831	ug/l	98
44) Trichloroethene	7.128	130	25296	18.275	ug/l	93
45) 1,2-Dichloropropane	7.433	63	21781	18.674	ug/l	100
46) Dibromomethane	7.586	93	15205	18.840	ug/l	99
47) Bromodichloromethane	7.823	83	29883	19.335	ug/l	98
48) Methyl methacrylate	7.695	41	26627	19.430	ug/l	97
49) 1,4-Dioxane	7.683	88	8748	370.437	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	173383	103.447	ug/l	99
52) Toluene	8.720	92	55611	19.053	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	28981	18.204	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	32627	19.117	ug/l	98
55) 1,1,2-Trichloroethane	9.152	97	22282	20.166	ug/l	95
56) Ethyl methacrylate	9.116	69	34815	19.877	ug/l	97
57) 1,3-Dichloropropane	9.311	76	38019	20.037	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	80219	89.232	ug/l	99
59) 2-Hexanone	9.433	43	126837	103.462	ug/l	99
60) Dibromochloromethane	9.524	129	22728	19.991	ug/l	99
61) 1,2-Dibromoethane	9.610	107	23010	19.710	ug/l	98
64) Tetrachloroethene	9.274	164	26344	17.346	ug/l	96
65) Chlorobenzene	10.079	112	58402	18.675	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	20936	18.198	ug/l	99
67) Ethyl Benzene	10.195	91	105436	19.049	ug/l	99
68) m/p-Xylenes	10.305	106	80618	37.856	ug/l	98
69) o-Xylene	10.640	106	39633	18.993	ug/l	97
70) Styrene	10.658	104	65916	19.476	ug/l	99
71) Bromoform	10.799	173	14770	17.820	ug/l #	99
73) Isopropylbenzene	10.963	105	104023	19.730	ug/l	99
74) N-amyl acetate	10.841	43	40736	19.337	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	27484	19.209	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	33584	22.921	ug/l	91
77) Bromobenzene	11.201	156	24156	19.275	ug/l	98
78) n-propylbenzene	11.305	91	116292	19.896	ug/l	99
79) 2-Chlorotoluene	11.366	91	70665	19.567	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	85216	19.638	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	7296	17.905	ug/l	99
82) 4-Chlorotoluene	11.457	91	80037	19.755	ug/l	99
83) tert-Butylbenzene	11.713	119	81654	19.743	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	85701	19.581	ug/l	98
85) sec-Butylbenzene	11.890	105	102716	19.593	ug/l	99
86) p-Isopropyltoluene	12.012	119	87252	19.646	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	43269	18.842	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	43300	18.697	ug/l	98
89) n-Butylbenzene	12.335	91	70329	19.355	ug/l	99
90) Hexachloroethane	12.542	117	12309	19.176	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	41784	18.857	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	6268	17.814	ug/l	97
93) 1,2,4-Trichlorobenzene	13.591	180	25105	17.953	ug/l	99
94) Hexachlorobutadiene	13.725	225	10803	18.264	ug/l	98
95) Naphthalene	13.780	128	90958	18.569	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	25406	17.942	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026888.D
Acq On : 09 Feb 2022 23:52
Operator : JC/MD
Sample : VX0209WBS02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Feb 10 01:13:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026888.D
 Acq On : 09 Feb 2022 23:52
 Operator : JC/MD
 Sample : VX0209WBS02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :


Quant Time: Feb 10 01:13:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:168 Resp: 100419
Ion Ratio Lower Upper
168 100
99 60.9 50.3 75.5






#2
Dichlorodifluoromethane
Concen: 19.656 ug/l
RT: 1.172 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 85 Resp: 21463
Ion Ratio Lower Upper
85 100
87 32.6 16.4 49.1



Abundance Scan 34 (1.294 min): VX026805.D\data.ms (-29)



#3

Chloromethane

Concen: 20.681 ug/l

RT: 1.294 min Scan# 34

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 50 Resp: 21007

Ion Ratio Lower Upper

50 100

52 33.8 26.1 39.1



#4

Vinyl Chloride

Concen: 20.426 ug/l

RT: 1.380 min Scan# 48

Delta R.T. -0.000 min

Lab File: VX026888.D


Acq: 09 Feb 2022 23:52

Tgt Ion: 62 Resp: 21720

Ion Ratio Lower Upper

62 100

64 32.0 26.2 39.2









#5
Bromomethane
Concen: 22.442 ug/l
RT: 1.617 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 10126
Ion Ratio Lower Upper
94 100
96 94.8 75.9 113.9



#6
Chloroethane
Concen: 19.316 ug/l
RT: 1.690 min Scan# 99
Delta R.T. -0.001 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 64 Resp: 12304
Ion Ratio Lower Upper
64 100
66 33.0 25.4 38.2









#7
Trichlorofluoromethane
Concen: 19.548 ug/l
RT: 1.892 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 30589
Ion Ratio Lower Upper
101 100
103 65.4 51.3 76.9



#8
Diethyl Ether
Concen: 19.360 ug/l
RT: 2.135 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 74 Resp: 11222
Ion Ratio Lower Upper
74 100
45 102.3 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.136 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:	101	Resp:	19628
Ion	Ratio	Lower	Upper
101	100		
85	43.2	35.3	52.9
151	77.1	61.4	92.2



#10
Methyl Iodide
Concen: 18.320 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion:	142	Resp:	23935
Ion	Ratio	Lower	Upper
142	100		
127	45.8	38.1	57.1
141	14.1	11.2	16.8





#11

Tert butyl alcohol

Concen: 88.302 ug/l

RT: 2.977 min Scan# 31

Delta R.T. -0.024 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 19229

Ion Ratio Lower Upper

59 100

57 10.2 7.2 10.8



#12

1,1-Dichloroethene

Concen: 18.890 ug/l

RT: 2.324 min Scan# 203

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52




Tgt Ion: 96 Resp: 19413


Ion Ratio Lower Upper

96 100

61 162.4 138.5 207.7


98 62.2 51.8 77.6








#13
Acrolein
Concen: 86.651 ug/l
RT: 2.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 17438
Ion Ratio Lower Upper
56 100
55 69.2 56.9 85.3



#14
Allyl chloride
Concen: 19.313 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 41 Resp: 35158
Ion Ratio Lower Upper
41 100
39 63.2 54.2 81.4
76 29.7 25.2 37.8









#15
Acrylonitrile
Concen: 101.644 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 63429
Ion Ratio Lower Upper
53 100
52 82.0 66.6 100.0
51 35.5 29.5 44.3



#16
Acetone
Concen: 100.473 ug/l
RT: 2.385 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 43 Resp: 53194
Ion Ratio Lower Upper
43 100
58 30.8 23.0 34.6









#17
Carbon Disulfide
Concen: 17.844 ug/l
RT: 2.520 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 51914
Ion Ratio Lower Upper
76 100
78 9.4 7.5 11.3



#18
Methyl Acetate
Concen: 18.846 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 43 Resp: 28708
Ion Ratio Lower Upper
43 100
74 22.9 18.3 27.5








#19
Methyl tert-butyl Ether
Concen: 18.817 ug/l
RT: 3.117 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 68449
Ion Ratio Lower Upper
73 100
57 20.6 17.4 26.2



#20
Methylene Chloride
Concen: 19.274 ug/l
RT: 2.794 min Scan# 280
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 84 Resp: 22111
Ion Ratio Lower Upper
84 100
49 123.2 101.5 152.3
51 36.9 31.2 46.8
86 62.4 51.3 76.9





#21

trans-1,2-Dichloroethene

Concen: 19.262 ug/l

RT: 3.099 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 20831

Ion Ratio Lower Upper

96 100

61 137.3 112.9 169.3

98 64.9 49.5 74.3



#22

Diisopropyl ether

Concen: 19.433 ug/l

RT: 3.763 min Scan# 439

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Tgt Ion: 45 Resp: 71224




Ion Ratio Lower Upper

45 100

43 81.1 60.6 91.0

87 28.2 22.0 33.0

59 11.3 9.0 13.4









#23
 Vinyl Acetate
 Concen: 98.075 ug/l
 RT: 3.727 min Scan# 411
 Delta R.T. -0.000 min
 Lab File: VX026888.D
 Acq: 09 Feb 2022 23:52




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 302214
 Ion Ratio Lower Upper
 43 100
 86 10.6 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 19.133 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. -0.000 min
 Lab File: VX026888.D
 Acq: 09 Feb 2022 23:52

Tgt Ion: 63 Resp: 37945
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.7 11.1
 100 4.2 2.4 7.2









#25
2-Butanone
Concen: 102.614 ug/l
RT: 4.562 min Scan# 51
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 87530
Ion Ratio Lower Upper
43 100
72 25.3 20.4 30.6



#26
2,2-Dichloropropane
Concen: 18.944 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52


Tgt Ion: 77 Resp: 25384
Ion Ratio Lower Upper
77 100
97 23.4 11.0 33.0







#27
 cis-1,2-Dichloroethene
 Concen: 19.424 ug/l
 RT: 4.495 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: VX026888.D
 Acq: 09 Feb 2022 23:52

Instrument :
 MSVOA_X
 ClientSampleId :



Tgt Ion: 96 Resp: 24118




Ion	Ratio	Lower	Upper
96	100		
61	142.1	0.0	296.6
98	63.6	0.0	128.8




#28
 Bromochloromethane
 Concen: 20.606 ug/l
 RT: 4.903 min Scan# 626
 Delta R.T. -0.000 min
 Lab File: VX026888.D
 Acq: 09 Feb 2022 23:52

Tgt Ion: 49 Resp: 16940

Ion	Ratio	Lower	Upper
49	100		
129	1.1	0.0	4.2
130	77.9	62.4	93.6





#29

Tetrahydrofuran

Concen: 101.507 ug/l

RT: 5.007 min Scan# 643

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 58864

Ion Ratio Lower Upper

42 100

72 44.3 34.6 52.0

71 39.8 31.8 47.8



#30

Chloroform

Concen: 19.358 ug/l

RT: 5.098 min Scan# 658

Delta R.T. -0.000 min

Lab File: VX026888.D




Acq: 09 Feb 2022 23:52


Tgt Ion: 83 Resp: 39887

Ion Ratio Lower Upper

83 100


85 64.7 51.6 77.4








#31
Cyclohexane
Concen: 18.887 ug/l
RT: 5.476 min Scan# 712
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 56 Resp: 36175




Ion	Ratio	Lower	Upper
56	100		
69	29.8	25.8	38.6
84	80.5	68.9	103.3



#32
1,1,1-Trichloroethane
Concen: 18.690 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 97 Resp: 33953

Ion	Ratio	Lower	Upper
97	100		
99	65.8	51.3	76.9
61	43.6	35.0	52.6





#33

1,2-Dichloroethane-d4

Concen: 55.414 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 66798

Ion Ratio Lower Upper

65 100

67 51.7 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52




Tgt Ion: 114 Resp: 168293


Ion Ratio Lower Upper

114 100

63 21.0 0.0 42.8


88 16.5 0.0 33.0








#35
Dibromofluoromethane
Concen: 52.367 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:113 Resp: 52506
Ion Ratio Lower Upper
113 100
111 101.3 81.2 121.8
192 19.3 15.0 22.4



#36
1,1-Dichloropropene
Concen: 18.660 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 75 Resp: 29515
Ion Ratio Lower Upper
75 100
110 35.2 17.2 51.6
77 30.6 24.6 36.8





#37

Ethyl Acetate

Concen: 18.910 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min


Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 43 Resp: 33992

Ion Ratio Lower Upper

43 100

61 13.5 10.6 16.0

70 11.0 8.5 12.7



#38

Carbon Tetrachloride

Concen: 18.215 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52




Tgt Ion: 117 Resp: 30213


Ion Ratio Lower Upper

117 100

119 99.0 74.7 112.1

121 30.0 25.4 38.0





#39

Methylcyclohexane

Concen: 18.456 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 35800

Ion Ratio Lower Upper

83 100

55 78.1 63.6 95.4

98 45.9 38.4 57.6



#40

Benzene

Concen: 18.544 ug/l

RT: 6.043 min Scan# 813

Delta R.T. -0.000 min

Lab File: VX026888.D




Acq: 09 Feb 2022 23:52


Tgt Ion: 78 Resp: 85489

Ion Ratio Lower Upper

78 100


77 23.8 18.9 28.3








#41
Methacrylonitrile
Concen: 19.156 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41 Resp: 18663
Ion Ratio Lower Upper
41 100
39 55.7 45.4 68.0
67 68.3 53.2 79.8
52 29.6 23.0 34.6



#42
1,2-Dichloroethane
Concen: 19.698 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52


Tgt Ion: 62 Resp: 31999
Ion Ratio Lower Upper
62 100
98 9.6 0.0 17.8








#43
Isopropyl Acetate
Concen: 18.831 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 52082
Ion Ratio Lower Upper
43 100
61 19.9 15.4 23.0
87 13.4 10.2 15.2



#44
Trichloroethene
Concen: 18.275 ug/l
RT: 7.128 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion:130 Resp: 25296
Ion Ratio Lower Upper
130 100
95 94.5 0.0 202.6





#45

1,2-Dichloropropane

Concen: 18.674 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 21781

Ion Ratio Lower Upper

63 100

65 29.9 23.8 35.6



#46

Dibromomethane

Concen: 18.840 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52




Tgt Ion: 93 Resp: 15205


Ion Ratio Lower Upper

93 100

95 82.3 66.1 99.1

174 100.6 78.7 118.1





#47

Bromodichloromethane

Concen: 19.335 ug/l

RT: 7.823 min Scan# 1105

Delta R.T. -0.000 min

Lab File: VX026888.D





Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 83	Resp: 29883
Ion Ratio	Lower Upper
83 100	
85 62.7	51.4 77.0
127 8.2	6.9 10.3



#48

Methyl methacrylate

Concen: 19.430 ug/l




RT: 7.695 min Scan# 1084

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Tgt Ion: 41	Resp: 26627
Ion Ratio	Lower Upper
41 100	
69 79.5	62.2 93.4
39 53.8	45.4 68.0





#49

1,4-Dioxane

Concen: 370.437 ug/l

RT: 7.683 min Scan# 1081

Delta R.T. 0.006 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 8748

Ion Ratio Lower Upper

88 100

43 36.8 29.9 44.9

58 69.6 57.8 86.6



#50

Toluene-d8

Concen: 54.401 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX026888.D




Acq: 09 Feb 2022 23:52

Tgt Ion: 98 Resp: 200221

Ion Ratio Lower Upper

98 100

100 66.5 53.2 79.8





#51

4-Methyl-2-Pentanone

Concen: 103.447 ug/l

RT: 8.573 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 173383

Ion Ratio Lower Upper

43 100

58 38.5 30.5 45.7



#52

Toluene

Concen: 19.053 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX026888.D




Acq: 09 Feb 2022 23:52


Tgt Ion: 92 Resp: 55611

Ion Ratio Lower Upper

92 100

91 170.6 135.7 203.5









#53
 t-1,3-Dichloropropene
 Concen: 18.204 ug/l
 RT: 8.982 min Scan# 1295
 Delta R.T. -0.000 min
 Lab File: VX026888.D
 Acq: 09 Feb 2022 23:52




Instrument :
 MSVOA_X
 ClientSampleId :

Tgt Ion: 75 Resp: 28981
 Ion Ratio Lower Upper
 75 100
 77 31.1 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 19.117 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: VX026888.D
 Acq: 09 Feb 2022 23:52

Tgt Ion: 75 Resp: 32627
 Ion Ratio Lower Upper
 75 100
 77 32.5 26.2 39.2
 39 52.3 43.5 65.3









#55
1,1,2-Trichloroethane
Concen: 20.166 ug/l
RT: 9.152 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 97 Resp: 22282




Ion	Ratio	Lower	Upper
97	100		
83	80.8	68.4	102.6
85	51.9	44.0	66.0
99	59.2	49.8	74.6




#56
Ethyl methacrylate
Concen: 19.877 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 69 Resp: 34815

Ion	Ratio	Lower	Upper
69	100		
41	69.1	57.2	85.8
39	37.0	30.9	46.3





#57

1,3-Dichloropropane

Concen: 20.037 ug/l

RT: 9.311 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 38019

Ion Ratio Lower Upper

76 100

78 32.1 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 89.232 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX026888.D




Acq: 09 Feb 2022 23:52

Tgt Ion: 63 Resp: 80219

Ion Ratio Lower Upper

63 100

106 29.2 23.0 34.6





#59

2-Hexanone

Concen: 103.462 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 126837

Ion Ratio Lower Upper

43 100

58 53.6 26.4 79.2



#60

Dibromochloromethane

Concen: 19.991 ug/l

RT: 9.524 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX026888.D




Acq: 09 Feb 2022 23:52


Tgt Ion: 129 Resp: 22728

Ion Ratio Lower Upper

129 100

127 76.1 38.5 115.5





#61

1,2-Dibromoethane

Concen: 19.710 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:107 Resp: 23010

Ion Ratio Lower Upper

107 100

109 96.1 75.5 113.3



#62

4-Bromofluorobenzene

Concen: 55.691 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52




Tgt Ion: 95 Resp: 72163

Ion Ratio Lower Upper

95 100

174 75.7 0.0 148.6


176 72.8 0.0 141.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 151219
Ion Ratio Lower Upper
117 100
82 55.6 45.2 67.8
119 32.4 26.2 39.2



#64
Tetrachloroethene
Concen: 17.346 ug/l
RT: 9.274 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion:164 Resp: 26344
Ion Ratio Lower Upper
164 100
166 131.7 101.8 152.6
129 96.5 83.3 124.9
131 96.4 78.0 117.0









#65
Chlorobenzene
Concen: 18.675 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 58402
Ion Ratio Lower Upper
112 100
114 31.6 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 18.198 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion:131 Resp: 20936
Ion Ratio Lower Upper
131 100
133 96.2 47.9 143.8
119 64.8 31.7 95.1





#67

Ethyl Benzene

Concen: 19.049 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 105436

Ion Ratio Lower Upper

91 100

106 31.3 24.7 37.1



#68

m/p-Xylenes

Concen: 37.856 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. -0.000 min

Lab File: VX026888.D




Acq: 09 Feb 2022 23:52


Tgt Ion: 106 Resp: 80618

Ion Ratio Lower Upper

106 100

91 203.2 165.3 247.9









#69
o-Xylene
Concen: 18.993 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 39633
Ion Ratio Lower Upper
106 100
91 215.1 109.7 329.0



#70
Styrene
Concen: 19.476 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion:104 Resp: 65916
Ion Ratio Lower Upper
104 100
78 51.1 42.5 63.7
103 54.2 43.3 64.9









#71
Bromoform
Concen: 17.820 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:173 Resp: 14770
Ion Ratio Lower Upper
173 100
175 47.8 24.3 72.9
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion:152 Resp: 66935
Ion Ratio Lower Upper
152 100
115 69.7 44.2 132.6
150 161.9 0.0 348.2









#73
Isopropylbenzene
Concen: 19.730 ug/l
RT: 10.963 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:105 Resp: 104023
Ion Ratio Lower Upper
105 100
120 26.7 13.1 39.1



#74
N-ethyl acetate
Concen: 19.337 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.006 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 43 Resp: 40736
Ion Ratio Lower Upper
43 100
70 43.0 32.9 49.3
55 24.6 18.3 27.5
61 24.8 18.9 28.3









#75
1,1,2,2-Tetrachloroethane
Concen: 19.209 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 27484
Ion Ratio Lower Upper
83 100
131 10.4 5.3 15.9
85 63.3 32.0 96.2



#76
1,2,3-Trichloropropane
Concen: 22.921 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52


Tgt Ion: 75 Resp: 33584
Ion Ratio Lower Upper
75 100
77 31.0 18.3 54.8








#77
Bromobenzene
Concen: 19.275 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:156 Resp: 24156
Ion Ratio Lower Upper
156 100
77 164.8 84.5 253.4
158 95.7 48.1 144.4



#78
n-propylbenzene
Concen: 19.896 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 91 Resp: 116292
Ion Ratio Lower Upper
91 100
120 23.4 11.6 34.7









#79
2-Chlorotoluene
Concen: 19.567 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 91 Resp: 70665
Ion Ratio Lower Upper
91 100
126 33.6 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 19.638 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52


Tgt Ion: 105 Resp: 85216
Ion Ratio Lower Upper
105 100
120 48.4 24.0 72.0








#81
trans-1,4-Dichloro-2-butene
Concen: 17.905 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 75 Resp: 7296
Ion Ratio Lower Upper
75 100
53 117.8 94.0 141.0
89 47.2 37.0 55.4



#82
4-Chlorotoluene
Concen: 19.755 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52


Tgt Ion: 91 Resp: 80037
Ion Ratio Lower Upper
91 100
126 28.8 14.1 42.4







#83
tert-Butylbenzene
Concen: 19.743 ug/l
RT: 11.713 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:119	Resp:	81654
Ion	Ratio	Lower Upper
119	100	
91	61.5	31.4 94.0
134	23.3	11.8 35.3



#84
1,2,4-Trimethylbenzene
Concen: 19.581 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52


Tgt Ion:105	Resp:	85701
Ion	Ratio	Lower Upper
105	100	
120	44.8	21.9 65.5







#85
 sec-Butylbenzene
 Concen: 19.593 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX026888.D
 Acq: 09 Feb 2022 23:52

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:105 Resp: 102716
 Ion Ratio Lower Upper
 105 100
 134 20.5 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 19.646 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX026888.D
 Acq: 09 Feb 2022 23:52

Tgt Ion:119 Resp: 87252
 Ion Ratio Lower Upper
 119 100
 134 25.5 13.0 39.0
 91 24.0 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 18.842 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 43269

Ion Ratio Lower Upper

146 100

111 40.8 21.3 63.7

148 64.8 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 18.697 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX026888.D

Acq: 09 Feb 2022 23:52




Tgt Ion:146 Resp: 43300


Ion Ratio Lower Upper

146 100

111 41.7 20.2 60.6

148 64.3 31.4 94.0









#89
n-Butylbenzene
Concen: 19.355 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 91 Resp: 70329




Ion	Ratio	Lower	Upper
91	100		
92	54.1	27.2	81.6
134	26.7	13.1	39.1




#90
Hexachloroethane
Concen: 19.176 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion: 117 Resp: 12309


Ion	Ratio	Lower	Upper
117	100		
201	73.6	35.6	106.8








#91
1,2-Dichlorobenzene
Concen: 18.857 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:146 Resp: 41784
Ion Ratio Lower Upper
146 100
111 42.8 21.8 65.3
148 64.6 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 17.814 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52




Tgt Ion: 75 Resp: 6268
Ion Ratio Lower Upper
75 100
155 82.2 40.0 119.9
157 107.3 51.6 154.8








#93
 1,2,4-Trichlorobenzene
 Concen: 17.953 ug/l
 RT: 13.591 min Scan# 2051
 Delta R.T. -0.000 min
 Lab File: VX026888.D
 Acq: 09 Feb 2022 23:52


Instrument :
 MSVOA_X
 ClientSampleId :





Tgt Ion:180 Resp: 25105
 Ion Ratio Lower Upper
 180 100
 182 94.0 46.7 140.0
 145 32.7 16.3 48.8




#94
 Hexachlorobutadiene
 Concen: 18.264 ug/l
 RT: 13.725 min Scan# 2073
 Delta R.T. -0.000 min
 Lab File: VX026888.D
 Acq: 09 Feb 2022 23:52




Tgt Ion:225 Resp: 10803
 Ion Ratio Lower Upper
 225 100
 223 63.1 32.0 96.0
 227 62.0 32.3 96.9







#95
Naphthalene
Concen: 18.569 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 90958
Ion Ratio Lower Upper
128 100
127 12.6 10.3 15.5
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 17.942 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX026888.D
Acq: 09 Feb 2022 23:52

Tgt Ion:180 Resp: 25406
Ion Ratio Lower Upper
180 100
182 97.9 47.2 141.6
145 33.9 17.5 52.5

