

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026889.D
 Acq On : 10 Feb 2022 00:15
 Operator : JC/MD
 Sample : VX0209WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 10 01:13:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	95329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	158370	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	143009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	63876	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	62418	54.546	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.100%	
35) Dibromofluoromethane	5.391	113	48890	51.816	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.640%	
50) Toluene-d8	8.653	98	188717	54.488	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.980%	
62) 4-Bromofluorobenzene	11.079	95	68190	55.922	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.840%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	22148	21.367	ug/l	98
3) Chloromethane	1.294	50	21259	22.046	ug/l	98
4) Vinyl Chloride	1.380	62	21971	21.766	ug/l	98
5) Bromomethane	1.618	94	10008	23.364	ug/l	99
6) Chloroethane	1.691	64	12982	21.627	ug/l	99
7) Trichlorofluoromethane	1.892	101	31000	20.869	ug/l	97
8) Diethyl Ether	2.136	74	11616	21.110	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.337	101	20052	20.593	ug/l	98
10) Methyl Iodide	2.459	142	25609	20.361	ug/l	98
11) Tert butyl alcohol	2.977	59	19567	94.651	ug/l	95
12) 1,1-Dichloroethene	2.325	96	19740	20.233	ug/l	93
13) Acrolein	2.239	56	19283	100.936	ug/l	99
14) Allyl chloride	2.672	41	36160	20.924	ug/l	95
15) Acrylonitrile	3.069	53	63117	106.544	ug/l	100
16) Acetone	2.386	43	53192	105.834	ug/l	97
17) Carbon Disulfide	2.520	76	54104	19.590	ug/l	99
18) Methyl Acetate	2.709	43	28937	20.011	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	70167	20.319	ug/l	99
20) Methylene Chloride	2.794	84	22577	20.731	ug/l	99
21) trans-1,2-Dichloroethene	3.099	96	21402	20.847	ug/l	95
22) Diisopropyl ether	3.764	45	72840	20.936	ug/l	100
23) Vinyl Acetate	3.727	43	309141	105.680	ug/l	98
24) 1,1-Dichloroethane	3.617	63	39243	20.844	ug/l	99
25) 2-Butanone	4.562	43	87540	108.105	ug/l	98
26) 2,2-Dichloropropane	4.483	77	26870	21.124	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	25138	21.326	ug/l	96
28) Bromochloromethane	4.904	49	16423	21.044	ug/l	98
29) Tetrahydrofuran	5.007	42	57314	104.111	ug/l	98
30) Chloroform	5.099	83	40353	20.630	ug/l	96
31) Cyclohexane	5.477	56	37212	20.466	ug/l	94
32) 1,1,1-Trichloroethane	5.391	97	35648	20.670	ug/l	99
36) 1,1-Dichloropropene	5.702	75	30419	20.436	ug/l	98
37) Ethyl Acetate	4.715	43	33209	19.632	ug/l	99
38) Carbon Tetrachloride	5.684	117	31291	20.047	ug/l	96
39) Methylcyclohexane	7.385	83	36771	20.144	ug/l	98
40) Benzene	6.044	78	88650	20.435	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026889.D
 Acq On : 10 Feb 2022 00:15
 Operator : JC/MD
 Sample : VX0209WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 10 01:13:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	19261	21.008	ug/l	96
42) 1,2-Dichloroethane	6.092	62	32818	21.468	ug/l	99
43) Isopropyl Acetate	6.342	43	53439	20.532	ug/l	99
44) Trichloroethene	7.129	130	26102	20.039	ug/l	97
45) 1,2-Dichloropropane	7.434	63	22668	20.652	ug/l	98
46) Dibromomethane	7.586	93	15744	20.730	ug/l	98
47) Bromodichloromethane	7.824	83	31074	21.365	ug/l	99
48) Methyl methacrylate	7.696	41	26648	20.663	ug/l	97
49) 1,4-Dioxane	7.684	88	8963	403.322	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	173214	109.822	ug/l	99
52) Toluene	8.720	92	57448	20.916	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	30479	20.185	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	34376	21.404	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	22687	21.819	ug/l	99
56) Ethyl methacrylate	9.116	69	35060	21.271	ug/l	98
57) 1,3-Dichloropropane	9.311	76	38921	21.798	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	84705	100.125	ug/l	99
59) 2-Hexanone	9.433	43	127061	110.139	ug/l	98
60) Dibromochloromethane	9.525	129	23293	21.771	ug/l	100
61) 1,2-Dibromoethane	9.610	107	24065	21.905	ug/l	98
64) Tetrachloroethene	9.275	164	27409	19.083	ug/l	97
65) Chlorobenzene	10.079	112	61122	20.666	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	21802	20.039	ug/l	99
67) Ethyl Benzene	10.195	91	109283	20.877	ug/l	100
68) m/p-Xylenes	10.305	106	83350	41.386	ug/l	97
69) o-Xylene	10.640	106	41038	20.795	ug/l	96
70) Styrene	10.659	104	67166	20.985	ug/l	98
71) Bromoform	10.799	173	15211	19.240	ug/l #	98
73) Isopropylbenzene	10.963	105	107644	21.395	ug/l	99
74) N-amyl acetate	10.842	43	41236	20.512	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.213	83	28132	20.604	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	33513	23.967	ug/l	95
77) Bromobenzene	11.201	156	24834	20.765	ug/l	97
78) n-propylbenzene	11.305	91	119404	21.407	ug/l	99
79) 2-Chlorotoluene	11.366	91	72436	21.018	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	88323	21.329	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	7322	18.655	ug/l	97
82) 4-Chlorotoluene	11.457	91	82695	21.389	ug/l	98
83) tert-Butylbenzene	11.713	119	83258	21.095	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	88471	21.181	ug/l	100
85) sec-Butylbenzene	11.890	105	106180	21.224	ug/l	100
86) p-Isopropyltoluene	12.012	119	89533	21.125	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	44505	20.308	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	44546	20.156	ug/l	99
89) n-Butylbenzene	12.335	91	72855	21.010	ug/l	98
90) Hexachloroethane	12.542	117	12790	20.880	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	43413	20.531	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	6387	19.022	ug/l	95
93) 1,2,4-Trichlorobenzene	13.591	180	25949	19.445	ug/l	98
94) Hexachlorobutadiene	13.725	225	11224	19.884	ug/l	98
95) Naphthalene	13.774	128	93194	19.937	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	26525	19.630	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026889.D
Acq On : 10 Feb 2022 00:15
Operator : JC/MD
Sample : VX0209WBSD02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Feb 10 01:13:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026889.D
 Acq On : 10 Feb 2022 00:15
 Operator : JC/MD
 Sample : VX0209WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Feb 10 01:13:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:168 Resp: 95329
Ion Ratio Lower Upper
168 100
99 60.7 50.3 75.5






#2
Dichlorodifluoromethane
Concen: 21.367 ug/l
RT: 1.173 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 85 Resp: 22148
Ion Ratio Lower Upper
85 100
87 31.6 16.4 49.1



Abundance Scan 34 (1.294 min): VX026805.D\data.ms (-29)



#3

Chloromethane

Concen: 22.046 ug/l

RT: 1.294 min Scan# 34

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X


ClientSampleId :

Tgt Ion: 50 Resp: 21259




Ion Ratio Lower Upper

50 100

52 31.6 26.1 39.1



Abundance Scan 48 (1.380 min): VX026805.D\data.ms (-42)



#4

Vinyl Chloride

Concen: 21.766 ug/l

RT: 1.380 min Scan# 48

Delta R.T. 0.000 min

Lab File: VX026889.D


Acq: 10 Feb 2022 00:15

Tgt Ion: 62 Resp: 21971

Ion Ratio Lower Upper

62 100

64 33.7 26.2 39.2









#5
Bromomethane
Concen: 23.364 ug/l
RT: 1.618 min Scan# 81
Delta R.T. 0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 10008
Ion Ratio Lower Upper
94 100
96 94.2 75.9 113.9



#6
Chloroethane
Concen: 21.627 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 64 Resp: 12982
Ion Ratio Lower Upper
64 100
66 32.1 25.4 38.2









#7
 Trichlorofluoromethane
 Concen: 20.869 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX026889.D
 Acq: 10 Feb 2022 00:15




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:101 Resp: 31000
 Ion Ratio Lower Upper
 101 100
 103 66.2 51.3 76.9



#8
 Diethyl Ether
 Concen: 21.110 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX026889.D
 Acq: 10 Feb 2022 00:15

Tgt Ion: 74 Resp: 11616
 Ion Ratio Lower Upper
 74 100
 45 97.6 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.593 ug/l
RT: 2.337 min Scan# 205
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:	101	Resp:	20052
Ion	Ratio	Lower	Upper
101	100		
85	42.4	35.3	52.9
151	78.8	61.4	92.2



#10
Methyl Iodide
Concen: 20.361 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion:	142	Resp:	25609
Ion	Ratio	Lower	Upper
142	100		
127	46.0	38.1	57.1
141	14.2	11.2	16.8








#11
Tert butyl alcohol
Concen: 94.651 ug/l
RT: 2.977 min Scan# 31
Delta R.T. -0.024 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 59 Resp: 19567
Ion Ratio Lower Upper
59 100
57 10.8 7.2 10.8



#12
1,1-Dichloroethene
Concen: 20.233 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 96 Resp: 19740
Ion Ratio Lower Upper
96 100
61 161.5 138.5 207.7
98 62.5 51.8 77.6





#13

Acrolein

Concen: 100.936 ug/l

RT: 2.239 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 19283

Ion Ratio Lower Upper

56 100

55 70.4 56.9 85.3



#14

Allyl chloride

Concen: 20.924 ug/l

RT: 2.672 min Scan# 260

Delta R.T. 0.006 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion: 41 Resp: 36160


Ion Ratio Lower Upper

41 100

39 62.8 54.2 81.4


76 30.9 25.2 37.8








#15
Acrylonitrile
Concen: 106.544 ug/l
RT: 3.069 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 63117
Ion Ratio Lower Upper
53 100
52 83.3 66.6 100.0
51 36.1 29.5 44.3



#16
Acetone
Concen: 105.834 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 43 Resp: 53192
Ion Ratio Lower Upper
43 100
58 30.4 23.0 34.6









#17
Carbon Disulfide
Concen: 19.590 ug/l
RT: 2.520 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 54104
Ion Ratio Lower Upper
76 100
78 9.2 7.5 11.3



#18
Methyl Acetate
Concen: 20.011 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 43 Resp: 28937
Ion Ratio Lower Upper
43 100
74 23.6 18.3 27.5








#19
Methyl tert-butyl Ether
Concen: 20.319 ug/l
RT: 3.117 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 73 Resp: 70167
Ion Ratio Lower Upper
73 100
57 21.2 17.4 26.2



#20
Methylene Chloride
Concen: 20.731 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion: 84 Resp: 22577
Ion Ratio Lower Upper
84 100
49 124.3 101.5 152.3
51 38.7 31.2 46.8
86 63.9 51.3 76.9








#21
trans-1,2-Dichloroethene
Concen: 20.847 ug/l
RT: 3.099 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 96 Resp: 21402
Ion Ratio Lower Upper
96 100
61 133.2 112.9 169.3
98 62.7 49.5 74.3



#22
Diisopropyl ether
Concen: 20.936 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 45 Resp: 72840
Ion Ratio Lower Upper
45 100
43 75.5 60.6 91.0
87 27.6 22.0 33.0
59 11.7 9.0 13.4









#23
 Vinyl Acetate
 Concen: 105.680 ug/l
 RT: 3.727 min Scan# 411
 Delta R.T. 0.000 min
 Lab File: VX026889.D
 Acq: 10 Feb 2022 00:15




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 309141
 Ion Ratio Lower Upper
 43 100
 86 10.9 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 20.844 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX026889.D
 Acq: 10 Feb 2022 00:15

Tgt Ion: 63 Resp: 39243
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.7 11.1
 100 4.7 2.4 7.2





#25

2-Butanone

Concen: 108.105 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 87540

Ion Ratio Lower Upper

43 100

72 24.5 20.4 30.6



#26

2,2-Dichloropropane

Concen: 21.124 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX026889.D




Acq: 10 Feb 2022 00:15


Tgt Ion: 77 Resp: 26870

Ion Ratio Lower Upper

77 100

97 23.2 11.0 33.0





#27

cis-1,2-Dichloroethene

Concen: 21.326 ug/l

RT: 4.495 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 25138

Ion Ratio Lower Upper

96 100

61 141.3 0.0 296.6

98 64.8 0.0 128.8



#28

Bromochloromethane

Concen: 21.044 ug/l

RT: 4.904 min Scan# 626

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion: 49 Resp: 16423


Ion Ratio Lower Upper

49 100

129 1.5 0.0 4.2


130 79.7 62.4 93.6








#29
Tetrahydrofuran
Concen: 104.111 ug/l
RT: 5.007 min Scan# 643
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 42 Resp: 57314




Ion	Ratio	Lower	Upper
42	100		
72	44.6	34.6	52.0
71	41.5	31.8	47.8




#30
Chloroform
Concen: 20.630 ug/l
RT: 5.099 min Scan# 658
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 83 Resp: 40353


Ion	Ratio	Lower	Upper
83	100		
85	67.4	51.6	77.4








#31
Cyclohexane
Concen: 20.466 ug/l
RT: 5.477 min Scan# 720
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 56 Resp: 37212




Ion	Ratio	Lower	Upper
56	100		
69	30.3	25.8	38.6
84	92.8	68.9	103.3



#32
1,1,1-Trichloroethane
Concen: 20.670 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 97 Resp: 35648

Ion	Ratio	Lower	Upper
97	100		
99	65.0	51.3	76.9
61	43.8	35.0	52.6





#33

1,2-Dichloroethane-d4

Concen: 54.546 ug/l

RT: 5.964 min Scan# 800

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 62418

Ion Ratio Lower Upper

65 100

67 51.7 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion: 114 Resp: 158370


Ion Ratio Lower Upper

114 100

63 20.2 0.0 42.8


88 15.2 0.0 33.0








#35
Dibromofluoromethane
Concen: 51.816 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:113 Resp: 48890
Ion Ratio Lower Upper
113 100
111 103.7 81.2 121.8
192 19.6 15.0 22.4



#36
1,1-Dichloropropene
Concen: 20.436 ug/l
RT: 5.702 min Scan# 757
Delta R.T. 0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 75 Resp: 30419
Ion Ratio Lower Upper
75 100
110 36.1 17.2 51.6
77 30.5 24.6 36.8





#37

Ethyl Acetate

Concen: 19.632 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min


Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 43 Resp: 33209

Ion Ratio Lower Upper

43 100

61 13.8 10.6 16.0

70 11.2 8.5 12.7



#38

Carbon Tetrachloride

Concen: 20.047 ug/l

RT: 5.684 min Scan# 754

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion: 117 Resp: 31291


Ion Ratio Lower Upper

117 100

119 96.7 74.7 112.1

121 29.1 25.4 38.0





#39

Methylcyclohexane

Concen: 20.144 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 36771

Ion Ratio Lower Upper

83 100

55 81.4 63.6 95.4

98 49.2 38.4 57.6



#40

Benzene

Concen: 20.435 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.000 min

Lab File: VX026889.D




Acq: 10 Feb 2022 00:15


Tgt Ion: 78 Resp: 88650

Ion Ratio Lower Upper

78 100


77 24.6 18.9 28.3








#41
Methacrylonitrile
Concen: 21.008 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41 Resp: 19261
Ion Ratio Lower Upper
41 100
39 52.4 45.4 68.0
67 68.7 53.2 79.8
52 27.6 23.0 34.6



#42
1,2-Dichloroethane
Concen: 21.468 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 62 Resp: 32818
Ion Ratio Lower Upper
62 100
98 9.3 0.0 17.8









#43
Isopropyl Acetate
Concen: 20.532 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 53439
Ion Ratio Lower Upper
43 100
61 19.4 15.4 23.0
87 13.1 10.2 15.2



#44
Trichloroethene
Concen: 20.039 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion:130 Resp: 26102
Ion Ratio Lower Upper
130 100
95 98.3 0.0 202.6





#45

1,2-Dichloropropane

Concen: 20.652 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 22668

Ion Ratio Lower Upper

63 100

65 30.8 23.8 35.6



#46

Dibromomethane

Concen: 20.730 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion: 93 Resp: 15744


Ion Ratio Lower Upper

93 100

95 83.2 66.1 99.1

174 101.3 78.7 118.1









#47
Bromodichloromethane
Concen: 21.365 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 83 Resp: 31074
Ion Ratio Lower Upper
83 100
85 63.7 51.4 77.0
127 9.4 6.9 10.3



#48
Methyl methacrylate
Concen: 20.663 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 41 Resp: 26648
Ion Ratio Lower Upper
41 100
69 81.2 62.2 93.4
39 55.2 45.4 68.0





#49

1,4-Dioxane

Concen: 403.322 ug/l

RT: 7.684 min Scan# 1082

Delta R.T. 0.006 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 8963

Ion Ratio Lower Upper

88 100

43 34.2 29.9 44.9

58 71.3 57.8 86.6



#50

Toluene-d8

Concen: 54.488 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. 0.000 min

Lab File: VX026889.D




Acq: 10 Feb 2022 00:15

Tgt Ion: 98 Resp: 188717

Ion Ratio Lower Upper

98 100

100 66.5 53.2 79.8





#51

4-Methyl-2-Pentanone

Concen: 109.822 ug/l

RT: 8.574 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 173214

Ion Ratio Lower Upper

43 100

58 38.8 30.5 45.7



#52

Toluene

Concen: 20.916 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. 0.000 min

Lab File: VX026889.D




Acq: 10 Feb 2022 00:15

Tgt Ion: 92 Resp: 57448

Ion Ratio Lower Upper

92 100

91 167.7 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 20.185 ug/l

RT: 8.982 min Scan# 1295

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 30479

Ion Ratio Lower Upper

75 100

77 31.4 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 21.404 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion: 75 Resp: 34376

Ion Ratio Lower Upper

75 100

77 31.6 26.2 39.2

39 50.6 43.5 65.3









#55
1,1,2-Trichloroethane
Concen: 21.819 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 97 Resp: 22687




Ion	Ratio	Lower	Upper
97	100		
83	87.3	68.4	102.6
85	54.4	44.0	66.0
99	63.2	49.8	74.6




#56
Ethyl methacrylate
Concen: 21.271 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 69 Resp: 35060

Ion	Ratio	Lower	Upper
69	100		
41	69.6	57.2	85.8
39	37.8	30.9	46.3









#57
1,3-Dichloropropane
Concen: 21.798 ug/l
RT: 9.311 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 38921
Ion Ratio Lower Upper
76 100
78 31.8 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 100.125 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 63 Resp: 84705
Ion Ratio Lower Upper
63 100
106 29.6 23.0 34.6





#59

2-Hexanone

Concen: 110.139 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 127061

Ion Ratio Lower Upper

43 100

58 53.9 26.4 79.2



#60

Dibromochloromethane

Concen: 21.771 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. 0.000 min

Lab File: VX026889.D




Acq: 10 Feb 2022 00:15


Tgt Ion: 129 Resp: 23293

Ion Ratio Lower Upper

129 100

127 77.1 38.5 115.5









#61
1,2-Dibromoethane
Concen: 21.905 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:107 Resp: 24065
Ion Ratio Lower Upper
107 100
109 92.4 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 55.922 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion: 95 Resp: 68190
Ion Ratio Lower Upper
95 100
174 76.2 0.0 148.6
176 72.9 0.0 141.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :




Tgt Ion:117 Resp: 143009

Ion	Ratio	Lower	Upper
117	100		
82	54.6	45.2	67.8
119	31.6	26.2	39.2






#64
Tetrachloroethene
Concen: 19.083 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15



Tgt Ion:164 Resp: 27409

Ion	Ratio	Lower	Upper
164	100		
166	125.2	101.8	152.6
129	98.1	83.3	124.9
131	94.4	78.0	117.0









#65
Chlorobenzene
Concen: 20.666 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 61122
Ion Ratio Lower Upper
112 100
114 31.4 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 20.039 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion:131 Resp: 21802
Ion Ratio Lower Upper
131 100
133 95.2 47.9 143.8
119 63.9 31.7 95.1





#67

Ethyl Benzene

Concen: 20.877 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 109283

Ion Ratio Lower Upper

91 100

106 30.7 24.7 37.1



#68

m/p-Xylenes

Concen: 41.386 ug/l

RT: 10.305 min Scan# 1512

Delta R.T. 0.000 min

Lab File: VX026889.D




Acq: 10 Feb 2022 00:15


Tgt Ion:106 Resp: 83350

Ion Ratio Lower Upper

106 100

91 202.5 165.3 247.9









#69
o-Xylene
Concen: 20.795 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 41038
Ion Ratio Lower Upper
106 100
91 213.7 109.7 329.0



#70
Styrene
Concen: 20.985 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion:104 Resp: 67166
Ion Ratio Lower Upper
104 100
78 50.1 42.5 63.7
103 54.1 43.3 64.9









#71
Bromoform
Concen: 19.240 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:173 Resp: 15211
Ion Ratio Lower Upper
173 100
175 47.6 24.3 72.9
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion:152 Resp: 63876
Ion Ratio Lower Upper
152 100
115 70.3 44.2 132.6
150 161.8 0.0 348.2









#73
Isopropylbenzene
Concen: 21.395 ug/l
RT: 10.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:105 Resp: 107644
Ion Ratio Lower Upper
105 100
120 26.5 13.1 39.1



#74
N-ethyl acetate
Concen: 20.512 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 43 Resp: 41236
Ion Ratio Lower Upper
43 100
70 43.9 32.9 49.3
55 25.1 18.3 27.5
61 24.3 18.9 28.3









#75
1,1,2,2-Tetrachloroethane
Concen: 20.604 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 28132
Ion Ratio Lower Upper
83 100
131 10.2 5.3 15.9
85 63.6 32.0 96.2



#76
1,2,3-Trichloropropane
Concen: 23.967 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion: 75 Resp: 33513
Ion Ratio Lower Upper
75 100
77 33.3 18.3 54.8








#77
Bromobenzene
Concen: 20.765 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:156 Resp: 24834
Ion Ratio Lower Upper
156 100
77 163.0 84.5 253.4
158 96.4 48.1 144.4



#78
n-propylbenzene
Concen: 21.407 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 91 Resp: 119404
Ion Ratio Lower Upper
91 100
120 23.4 11.6 34.7









#79
2-Chlorotoluene
Concen: 21.018 ug/l
RT: 11.366 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 91 Resp: 72436
Ion Ratio Lower Upper
91 100
126 33.3 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 21.329 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 105 Resp: 88323
Ion Ratio Lower Upper
105 100
120 49.1 24.0 72.0









#81
trans-1,4-Dichloro-2-butene
Concen: 18.655 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 7322
Ion Ratio Lower Upper
75 100
53 121.4 94.0 141.0
89 47.9 37.0 55.4



#82
4-Chlorotoluene
Concen: 21.389 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 91 Resp: 82695
Ion Ratio Lower Upper
91 100
126 29.2 14.1 42.4









#83
tert-Butylbenzene
Concen: 21.095 ug/l
RT: 11.713 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 83258
Ion Ratio Lower Upper
119 100
91 62.0 31.4 94.0
134 23.8 11.8 35.3



#84
1,2,4-Trimethylbenzene
Concen: 21.181 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion:105 Resp: 88471
Ion Ratio Lower Upper
105 100
120 43.8 21.9 65.5







#85
sec-Butylbenzene
Concen: 21.224 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:105 Resp: 106180
Ion Ratio Lower Upper
105 100
134 20.1 10.1 30.1



#86
p-Isopropyltoluene
Concen: 21.125 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion:119 Resp: 89533
Ion Ratio Lower Upper
119 100
134 25.8 13.0 39.0
91 24.0 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 20.308 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 44505

Ion Ratio Lower Upper

146 100

111 41.8 21.3 63.7

148 64.5 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 20.156 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX026889.D

Acq: 10 Feb 2022 00:15




Tgt Ion:146 Resp: 44546


Ion Ratio Lower Upper

146 100

111 41.2 20.2 60.6

148 63.7 31.4 94.0









#89
n-Butylbenzene
Concen: 21.010 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 72855
Ion Ratio Lower Upper
91 100
92 53.1 27.2 81.6
134 26.7 13.1 39.1



#90
Hexachloroethane
Concen: 20.880 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion: 117 Resp: 12790
Ion Ratio Lower Upper
117 100
201 74.0 35.6 106.8









#91
1,2-Dichlorobenzene
Concen: 20.531 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 43413
Ion Ratio Lower Upper
146 100
111 42.1 21.8 65.3
148 62.5 32.0 96.2



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.022 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Tgt Ion: 75 Resp: 6387
Ion Ratio Lower Upper
75 100
155 86.5 40.0 119.9
157 106.7 51.6 154.8








#93
1,2,4-Trichlorobenzene
Concen: 19.445 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:180 Resp: 25949
Ion Ratio Lower Upper
180 100
182 95.2 46.7 140.0
145 32.1 16.3 48.8




#94
Hexachlorobutadiene
Concen: 19.884 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15




Tgt Ion:225 Resp: 11224
Ion Ratio Lower Upper
225 100
223 62.2 32.0 96.0
227 62.6 32.3 96.9







#95
Naphthalene
Concen: 19.937 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.006 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 93194
Ion Ratio Lower Upper
128 100
127 12.6 10.3 15.5
129 10.7 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 19.630 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX026889.D
Acq: 10 Feb 2022 00:15

Tgt Ion:180 Resp: 26525
Ion Ratio Lower Upper
180 100
182 94.4 47.2 141.6
145 33.4 17.5 52.5

