

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026866.D
 Acq On : 09 Feb 2022 12:43
 Operator : JC/MD
 Sample : VX0209WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0209WBL01

Quant Time: Feb 10 01:04:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221520	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195208	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84630	52.961	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.920%	
35) Dibromofluoromethane	5.391	113	68665	52.028	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.060%	
50) Toluene-d8	8.653	98	255802	52.802	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.600%	
62) 4-Bromofluorobenzene	11.079	95	89685	52.583	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.160%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026866.D
Acq On : 09 Feb 2022 12:43
Operator : JC/MD
Sample : VX0209WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0209WBL01

Quant Time: Feb 10 01:04:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

