

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026871.D
 Acq On : 09 Feb 2022 16:11
 Operator : JC/MD
 Sample : N1429-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22-R-SDL

Quant Time: Feb 10 01:06:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	214498	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185991	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78597	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83306	54.391	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.780%	
35) Dibromofluoromethane	5.391	113	67471	52.797	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.600%	
50) Toluene-d8	8.653	98	244269	52.072	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.140%	
62) 4-Bromofluorobenzene	11.079	95	86674	52.481	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.960%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.380	62	45950	34.010	ug/l	98
21) trans-1,2-Dichloroethene	3.099	96	1385	1.008	ug/l	88
27) cis-1,2-Dichloroethene	4.495	96	86116	54.584	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026871.D
Acq On : 09 Feb 2022 16:11
Operator : JC/MD
Sample : N1429-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22-R-SDL

Quant Time: Feb 10 01:06:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

