

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026876.D
 Acq On : 09 Feb 2022 18:06
 Operator : JC/MD
 Sample : N1438-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22-R-D

Quant Time: Feb 10 01:08:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	217323	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194431	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81940	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83192	53.560	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.120%	
35) Dibromofluoromethane	5.391	113	67812	52.374	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.740%	
50) Toluene-d8	8.653	98	251360	52.888	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.780%	
62) 4-Bromofluorobenzene	11.079	95	91140	54.468	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.940%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.380	62	52719	38.476	ug/l	98
12) 1,1-Dichloroethene	2.325	96	1629	1.230	ug/l	85
16) Acetone	2.386	43	1977	2.898	ug/l	95
21) trans-1,2-Dichloroethene	3.099	96	6628	4.756	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	1008886	630.569	ug/l	88
39) Methylcyclohexane	7.385	83	1607	0.642	ug/l	96
40) Benzene	6.044	78	3219	0.541	ug/l	94
44) Trichloroethene	7.129	130	5070	2.836	ug/l	96
64) Tetrachloroethene	9.275	164	2002	1.025	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026876.D
Acq On : 09 Feb 2022 18:06
Operator : JC/MD
Sample : N1438-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22-R-D

Quant Time: Feb 10 01:08:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

