

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026893.D
 Acq On : 10 Feb 2022 01:47
 Operator : JC/MD
 Sample : N1447-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB020822

Quant Time: Feb 10 03:53:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	133740	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226644	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199837	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80986	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86677	53.990	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.980%	
35) Dibromofluoromethane	5.391	113	70544	52.243	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.480%	
50) Toluene-d8	8.653	98	262523	52.965	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.920%	
62) 4-Bromofluorobenzene	11.079	95	92149	52.806	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.620%	
Target Compounds						
16) Acetone	2.386	43	1692	2.400	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026893.D
Acq On : 10 Feb 2022 01:47
Operator : JC/MD
Sample : N1447-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB020822

Quant Time: Feb 10 03:53:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

