

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
 Data File : VX026894.D
 Acq On : 10 Feb 2022 02:10
 Operator : JC/MD
 Sample : N1447-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB020922

Quant Time: Feb 10 03:53:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 08 07:07:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206209	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189190	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82644	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79511	54.092	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.180%	
35) Dibromofluoromethane	5.397	113	64498	52.499	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.000%	
50) Toluene-d8	8.653	98	242374	53.745	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.500%	
62) 4-Bromofluorobenzene	11.079	95	92438	58.221	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 116.440%	
Target Compounds						
16) Acetone	2.386	43	2010	3.113	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX020922\
Data File : VX026894.D
Acq On : 10 Feb 2022 02:10
Operator : JC/MD
Sample : N1447-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB020922

Quant Time: Feb 10 03:53:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X020822W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 08 07:07:46 2022
Response via : Initial Calibration

